



Standard Test Methods for Measuring and Compensating for Transmittance of an Attenuating Medium Using Infrared Imaging Radiometers¹

This standard is issued under the fixed designation E 1897; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

^{e1} NOTE—Removed Footnote 2 in August 2002 since the organization is no longer in business.

1. Scope

1.1 These test methods cover procedures for measuring and compensating for transmittance when using an infrared imaging radiometer to measure the temperature of a specimen through an attenuating medium, such as a window, filter or atmosphere.

1.2 The values stated in SI units are to be regarded as the standard.

1.3 *These test methods may involve use of equipment and materials in the presence of heated or electrically-energized equipment, or both.*

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 *ASTM Standards:*

E 1316 Terminology for Nondestructive Examinations²

3. Terminology

3.1 *Definitions of Terms Specific to This Standard:*

3.1.1 *attenuating medium*—a semi-transparent solid, liquid or gas, such as a window, filter, external optics and/or an atmosphere that attenuates radiation.

3.1.2 *blackbody simulator*—a device with an emissivity close to 1.00 that can be heated or cooled to a stable temperature.

3.1.3 *filter*—a semi-transparent material that attenuates certain wavelengths of radiation.

3.1.4 *infrared thermographer*—the person using an infrared imaging radiometer.

3.1.5 *reflected temperature*—the temperature of the energy incident upon and reflected by the measurement surface of the specimen.

3.1.6 *window*—a semi-transparent material that separates conditioned and unconditioned atmospheres and attenuates certain wavelengths of radiation.

3.2 See also Terminology E 1316.

4. Summary of Test Method

4.1 Using the computer built into an infrared imaging radiometer, a method is given for measuring the transmittance of an attenuating medium.

4.2 Using the computer built into an infrared imaging radiometer, a method is given for compensating for errors when measuring the temperature of a specimen through an attenuating medium when the emissivity of the specimen and the transmittance of the attenuating medium are known.

4.3 Using the computer built into an infrared imaging radiometer, a method is given for measuring and compensating for unknown transmittance and emissivity errors when the specimen temperature is known.

5. Significance and Use

5.1 The transmittance of an attenuating medium can cause errors for an infrared thermographer using an infrared imaging radiometer to measure the temperature of a specimen through the medium. Three test methods are given for measuring and compensating for this error source.

5.1.1 A test method is given for measuring the transmittance of an attenuating medium.

5.1.2 A test method is given for compensating for errors when measuring the temperature of a specimen having a known emissivity through an attenuating medium with a known transmittance.

5.1.3 A test method is given for measuring and compensating for transmittance and emissivity errors when the specimen temperature is known.

5.2 These test methods can be used in the field or laboratory using commonly available materials.

¹ These test methods are under the jurisdiction of ASTM Committee E07 on Nondestructive Testing and is the direct responsibility of Subcommittee E07.10 on Emerging NDT Methods.

Current edition approved July 10, 1997. Published December 1997.

² *Annual Book of ASTM Standards*, Vol 03.03.

5.3 These test methods can be used with any infrared radiometers that have the required computer capabilities.

6. Interferences

6.1 *Test Method for Measuring the Transmittance of an Attenuating Medium:*

6.1.1 This test method requires a blackbody simulator with an emissivity of 0.95 or greater that is at least 20°C warmer than ambient temperature. Potential errors can be minimized by ensuring the stability of the temperature difference between the source and the ambient temperature during the test. Also, the transmittance measurement accuracy can be increased by increasing this temperature difference.

6.1.2 Errors can be minimized by ensuring that the temperature of the attenuating medium is as close as possible to the reflected temperature incident upon the specimen.

6.1.3 The compositions and thicknesses of attenuating media can vary within the same specimen. Errors can be minimized by measuring through the same portion of the specimen every time.

6.2 *Test Method for Compensating for Unknown Errors When Measuring the Temperature of a Specimen With a Known Emissivity Through an Attenuating Medium Having a Known Transmittance:*

6.2.1 Errors can be minimized by ensuring that the temperature of the attenuating medium is as close as possible to the reflected temperature incident upon the specimen.

6.3 The transmittance of an attenuating medium may be specific to the temperature of the medium and the spectral waveband of the radiometer used to make the measurement. Therefore, the temperature of the measured specimen and the spectral waveband of the radiometer used should be noted with the measured transmittance value.

6.4 The use of these test methods to compensate for the transmittance of an attenuating medium is valid only when measuring the temperatures of specimens that are opaque in the waveband of the infrared imaging radiometer.

7. Apparatus

7.1 A calibrated infrared imaging radiometer with a built-in computer that allows the infrared thermographer to input reflected temperatures and emissivity values.

7.2 The test method for measuring the transmittance of an attenuating medium requires a tripod or device to support the infrared imaging radiometer.

7.3 The test method for measuring the transmittance of an attenuating medium requires a high-emissivity source that is heated to a stable temperature at least 20°C above ambient temperature.

7.4 The test method for measuring and compensating for unknown transmittance and emissivity errors when the specimen temperature is known requires a calibrated thermometer to measure the temperature of the specimen.

8. Procedure

8.1 To measure the transmittance of an attenuating medium, use the following sequential steps:

8.1.1 Place the infrared imaging radiometer on the tripod or support device at the desired location and distance from the blackbody simulator.

8.1.2 Point the infrared imaging radiometer at the blackbody simulator and focus on a portion that has an emissivity of 0.95 or greater. Make sure that the blackbody simulator is at a stable temperature at least 20°C above the ambient temperature.

8.1.3 Use an appropriate infrared imaging radiometer measurement function (such as spot temperature, crosshairs or isotherm) to measure and compensate for the reflected temperature error incident upon the blackbody simulator.

8.1.4 With the imager's computer emittance control still set to 1.00, measure and record the apparent temperature of this same portion of the blackbody simulator.

8.1.5 Position the attenuating medium between the imager's detector or lens and the blackbody simulator.

8.1.6 Without moving the imager, adjust its computer's emissivity control until the imager's computer indicates the same temperature recorded in 8.1.4. The indicated "emissivity" value is the transmittance of the attenuating medium, at this blackbody simulator temperature and radiometer's spectral waveband.

8.1.7 Repeat 8.1.1 through 8.1.6 a minimum of three times and average the transmittance values to yield an average transmittance.

8.2 To measure the temperature of a specimen having a known emissivity through an attenuating medium having a known transmittance, use the following sequential steps.

8.2.1 Point the infrared imaging radiometer at the specimen and focus on the portion where the temperature is to be measured. Place the attenuating medium having a known transmittance between the imager's detector or lens and the specimen.

8.2.2 Instead of 1.00, enter the measured transmittance percentage of the attenuating medium under the imager's emissivity input (sometime referred to as "emittance" or "E"). Use an appropriate infrared imaging radiometer function (such as spot temperature, crosshairs or isotherm) to measure and compensate for the reflected temperature error incident upon the specimen.

8.2.3 Calculate the combined emissivity and transmittance correction value by multiplying the known emissivity of the specimen times the known transmittance of the attenuating medium. Record this combined correction value.

8.2.4 Repeat 8.2.1 through 8.2.3 a minimum of three times and average the values to yield an average combined correction value.

8.2.5 Compensate for emissivity and transmittance errors by entering the combined correction value recorded in 8.2.3 in the infrared radiometric imager's computer under the "emissivity" input.

8.3 To measure and compensate for unknown transmittance and emissivity errors when the specimen temperature is known, use the following sequential steps. Use an appropriate infrared imaging radiometer function (such as spot temperature, crosshairs or isotherm) to measure and compensate for the reflected temperature error incident upon the specimen.

8.3.1 Point the imager at the specimen and the portion of the specimen having an unknown emissivity. Place the attenuating medium(s) with an unknown transmittance(s) between the imager's detector or lens and the specimen.

8.3.2 With another thermometer, measure and record the temperature of the portion of the specimen delineated with the infrared imaging radiometer's measurement function.

8.3.3 With the imager still focused on the same portion of the specimen, adjust the imager's computer emissivity control until the indicated temperature is the same as the temperature just recorded in 8.3.2. Record this "emissivity" value which is the combined correction value for errors produced by the transmittance of the attenuating medium and the emissivity of the specimen, at this medium temperature and radiometer's spectral waveband.

8.3.4 Repeat 8.3.1 through 8.3.3 a minimum of three times and average the emissivity values to yield an average combined correction value.

8.3.5 Enter this average combined correction value in the imager computer under the emissivity input to compensate for errors produced by this attenuating medium when measuring a specimen with the same emissivity.

9. Precision and Bias

9.1 *Precision*—A measure of the precision of these test methods can be inferred from the results of the replicate tests specified in 8.1.7, 8.2.4, and 8.3.4.

9.2 *Bias*—The bias cannot be determined by these test methods because the values of transmittance are defined only in terms of the test methods.

10. Keywords

10.1 attenuating medium; blackbody simulator; emissivity; filter; imaging, infrared; infrared examination; infrared testing; infrared thermography; nondestructive testing; radiometry; reflected temperature; temperature compensation; transmittance; temperature measurement; window

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).