



Standard Test Method for Traces of Thiophene in Benzene by Spectrophotometry¹

This standard is issued under the fixed designation D 1685; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last approval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope*

1.1 This test method covers the determination of thiophene in benzene in which the thiophene concentration is between 0.1 and 250 mg/kg.

1.2 Contaminating materials that are darkened by sulfuric acid interfere and must be compensated for by a sample blank determination. Neither carbon disulfide in concentrations as high as 100 mg/kg nor water as high as the level of saturation will interfere.

1.3 Contaminating materials may occasionally cause the color development in this test method to be time dependent and may also contribute to spectral interferences at 589 nm. Consequently, test results indicating substantial thiophene, that cannot be verified by total sulfur analysis, should be considered suspect.

1.4 The following applies to all specified limits in this test method: for purposes of determining conformance with this test method, an observed value or a calculated value shall be round off “to the nearest unit” in the last right-hand digit used in expressing the specification limit, in accordance with the round-off method of Practice E 29.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For specific hazard statements, see Note 1, 8.1, and 9.1.

2. Referenced Documents

2.1 ASTM Standards:

D 1193 Specification for Reagent Water²

D 3437 Practice for Sampling and Handling Liquid Cyclic Products³

D 4790 Terminology of Aromatic Hydrocarbons and Related Chemicals³

¹ This test method is under the jurisdiction of ASTM Committee D16 on Aromatic Hydrocarbons and Related Chemicals and is the direct responsibility of Subcommittee D16.01 on Benzene, Toluene, Xylene, Cyclohexane, and Their Derivatives.

Current edition approved June 10, 2000. Published August 2000. Originally published as D 1685 – 58 T. Last previous edition D 1685 – 95.

² *Annual Book of ASTM Standards*, Vol 11.01.

³ *Annual Book of ASTM Standards*, Vol 06.04.

E 29 Practice for Using Significant Digits in Test Data to Determine Conformance with Specifications⁴

2.2 Other Document:

OSHA Regulations, 29 CFR, paragraphs 1910.1000 and 1910.1200⁵

3. Terminology

3.1 Definitions:

3.1.1 For definition of terms used in this test method see Terminology D 4790.

4. Summary of Test Method

4.1 Thiophene is reacted with isatin, under prescribed conditions, to form a colored compound. The compound is extracted into sulfuric acid, and the intensity of the color is measured spectrophotometrically. Thiophene concentration is obtained by correlation with knowns.

5. Significance and Use

5.1 This test method is suitable for setting specifications on benzene and for use as an internal quality control tool where benzene is either produced or used in a manufacturing process. It may also be used in development or research work involving benzene.

6. Apparatus

6.1 *Separatory Funnels*, 50, 250, 500, and 1000-mL, with glass stoppers.

6.2 *Spectrophotometer*—Any spectrophotometer may be used that is capable of repeatability of 0.005 absorbance units in the range from 0.1 to 1.4 absorbance and repeatability of wavelength of 1 nm in the region from 400 to 700 nm.

6.3 *Absorption Cells*, 1-cm, matched, glass or silica.

6.4 *Analytical Balance*.

6.5 *Pipets*, 1, 2, 5, and 10-mL.

6.6 *Graduated Cylinders*, 250 and 1000-mL.

6.7 *Volumetric Flasks*, glass-stoppered, 50, 100, 250, and 1000-mL.

⁴ *Annual Book of ASTM Standards*, Vol 14.02.

⁵ Available from Superintendent of Documents, U.S. Government Printing Office, Washington, DC 20402.

*A Summary of Changes section appears at the end of this standard.

6.8 *Filter Paper*, medium filter⁶ and rapid hardened.⁷

7. Reagents

7.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.⁸ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

7.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean Type IV reagent water conforming to Specification D 1193.

7.3 *Cadmium Chloride Solution* (20 g/L)—Dissolve 20 g of anhydrous cadmium chloride (CdCl_2) or 25 g of cadmium chloride hydrate ($\text{CdCl}_2 \cdot 2\frac{1}{2} \text{H}_2\text{O}$) in 200 mL of water and dilute to 1 L.

7.4 *Sulfuric Acid* (sp gr 1.84)—Concentrated H_2SO_4 .

7.5 *Benzene, Thiophene-Free*:

7.5.1 Wash 700 mL of benzene in a 1000-mL separatory funnel with successive 100-mL portions of concentrated sulfuric acid (H_2SO_4) to which has been added 5 mL of isatin solution, until the H_2SO_4 layer is light yellow or colorless. Wash the benzene with 100 mL of water and then twice with 100 mL of cadmium chloride solution (CdCl_2). Finally, wash with another 100-mL portion of water. Filter the benzene through medium filter paper into a storage bottle and tightly stopper.

7.5.2 Prepare 1400 mL of thiophene-free benzene. Measure the absorbance of this material by the procedure outlined in 11.2 and 12.2. The absorbance should be no greater than 0.01.

7.6 *Ferric Sulfate, Sulfuric Acid Solution*—Add 0.2 g of ferric sulfate ($\text{Fe}_2(\text{SO}_4)_3 \cdot 9\text{H}_2\text{O}$) together with 38 mL of water to a 1-L volumetric flask. Swirl to dissolve. Cautiously add about 100 mL of H_2SO_4 and swirl. Allow time for the heat of reaction to subside and dilute to volume with H_2SO_4 .

NOTE 1—**Precaution:** Protective clothing and goggles should be worn whenever H_2SO_4 is used.

7.7 *Isatin, Chloroform, Benzene Solution*—Add 0.5 g of isatin to 200 mL of chloroform. Heat, in a fume hood, to a temperature just below the boiling point (61°C) of chloroform and maintain for 5 min with stirring. Filter into a 250-mL volumetric flask through hardened rapid filter paper. Wash the filter paper with two 20-mL portions of thiophene-free benzene (from 7.5) eluting the washings into the volumetric flask. Dilute to volume with thiophene-free benzene.

7.8 *Thiophene*.⁹

⁶ This method is based on the use of Whatman 1.

⁷ This method is based on the use of Whatman 54.

⁸ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

⁹ Thiophene such as Eastman Catalog No. 1860 or equivalent has been found satisfactory for this purpose.

8. Hazards

8.1 Consult current OSHA regulations and supplier's Material Safety Data Sheets for all materials used in this test method.

9. Sampling

9.1 Sampling of benzene should follow safe rules in order to adhere to all safety precautions as outlined in the latest OSHA regulations. Refer to Practice D 3437 for proper sampling and handling of benzene.

10. Preparation of Reagent Blanks and Sample Blank

10.1 *Reagent Blank 1*—To a 50-mL separatory funnel pipet 5 mL of isatin solution and 10 mL of ferric sulfate-sulfuric acid solution. Stopper and shake for 2 min \pm 15 s. The shaking is accomplished by wrist action in a rocking motion through a 180° arc roughly once each second. Allow the two phases to separate and draw off the lower H_2SO_4 layer into a 50-mL volumetric flask. Add 10 mL of H_2SO_4 to the separatory funnel, stopper, and shake for 30 \pm 5 s. Again draw off the lower H_2SO_4 layer into the 50-mL volumetric flask containing the first extract. Dilute to volume with H_2SO_4 and mix. This blank is stable for 8 h and need not be repeated with each analysis during this period.

10.2 *Reagent Blank 2*—Into a 50-mL volumetric flask, pipet 10 mL of ferric sulfate-sulfuric acid solution and dilute to volume with H_2SO_4 . Stopper and mix. This blank is stable for 8 h and need not be repeated with each analysis during this period.

10.3 *Sample Blank*—Take a 100-mL portion of the CdCl_2 washed and filtered benzene sample (prepared in accordance with the procedure in 12.1). Transfer to a 250-mL separatory funnel. Add 10 mL of ferric sulfate-sulfuric acid solution, stopper, and shake for 2 min \pm 15 s. Allow the two phases to separate and draw off the lower H_2SO_4 layer into a 50-mL volumetric flask. Add 10 mL of H_2SO_4 to the separatory funnel and shake for 30 \pm 5 s. Again draw off the lower H_2SO_4 layer into the 50-mL volumetric flask containing the first extract. Dilute to volume and mix. Repeat with each specimen.

11. Preparation of Calibration Curves

11.1 Add approximately 0.2 g of thiophene, weighed to the nearest 0.0002 g to a 100-mL volumetric flask containing about 50 mL of thiophene-free benzene. Dilute to volume with thiophene-free benzene and mix. This is Solution 1. Pipet 1 mL of Solution 1 into a 100-mL volumetric flask, dilute to volume with thiophene-free benzene, stopper, and mix. This is Solution 2 and will contain approximately 20 μg of thiophene per millilitre.

11.2 Pipet 0, 1, 2, 5, 7, and 10-mL of Solution 2 into 100-mL volumetric flasks and dilute to volume with thiophene-free benzene. Transfer to 500-mL separatory funnels and follow the procedure in 12.2 and 12.3 for each concentration. Plot absorbance versus concentration in micrograms per millilitre.

11.3 Prepare Solution 3 containing approximately 40 μg of thiophene per millilitre by pipetting 2 mL of Solution 1 into a 100-mL volumetric flask and diluting to volume with

TABLE 1 Calculation Factors

Designation of Equation of Section 12	Cell Solutions		
	Solution of Sample	Versus	Solution of Reference
A	H ₂ SO ₄ -Fe ₂ (SO ₄) ₃ ; isatin after contact with sample, (see 12.2)		reagent blank 1 (see 10.1)
B	sample blank (see 10.3)		reagent blank 2 (see 10.2)

thiophene-free benzene. Follow the procedure in 11.2 to obtain the calibration curve for 50-mL specimens.

12. Procedure

12.1 To a 500-mL separatory funnel, add 250 mL of sample and 40 mL of cadmium chloride solution. Stopper and shake for approximately 30 s. Allow to settle and discard the aqueous layer. Filter the benzene layer through medium filter paper into a 250-mL graduated cylinder. Part of the filtered benzene is to be used for the sample blank. Proceed with preparation of the sample blank as described in 10.3. From the remaining filtered benzene, transfer 100 mL to 250-mL separatory funnel.

12.2 To the separatory funnel add 5 mL of isatin solution and 10 mL of ferric sulfate-sulfuric acid solution. Stopper and shake for 2 min ± 15 s. Allow the phases to separate and draw off the lower H₂SO₄ layer into a 50-mL volumetric flask. Add 10 mL of H₂SO₄ to the separatory funnel, stopper, and shake for 30 ± 5 s. Again draw off the lower H₂SO₄ layer into the 50-mL volumetric flask containing the first extract. Dilute to volume with H₂SO₄ and mix.

12.3 Measure the absorbance of this material at 589 nm in a 1-cm cell versus Reagent Blank 1 (10.1) in a matched 1-cm cell. Instrument conditions should be identical with those employed during calibration.

12.4 Determine the concentration of thiophene from the calibration curve (Section 11). If the absorbance is greater than 1.5, repeat the procedure using a 50-mL specimen instead of 100 mL. If the 50-mL specimen still gives absorbance above 1.5, then the specimen must be diluted with thiophene-free benzene before proceeding.

12.5 Determine the absorbance of the sample blank (10.3) at 589 nm using Reagent Blank 2 (10.2) as reference. Determine the apparent concentration of thiophene in the sample blank.

13. Calculation

13.1 Calculate the thiophene content of the sample in milligrams per kilogram as follows:

$$\text{Thiophene, mg/kg} = (A - B)F/d$$

where:

A = thiophene for sample determined from appropriate calibration curve (see Table 1), µg/mL,

B = apparent thiophene determined for sample blank from appropriate calibration curve (see Table 1), µg/mL,

F = dilution factor of sample, and

d = density of benzene at the temperature of the sample.

14. Precision and Bias

14.1 The data given in Table 2 should be used for judging the

TABLE 2 Precision Limits

Thiophene Concentration, Range, mg/kg	Repeatability, Percent of Value	Reproducibility, Percent of Value
100 to 250	11.4	15
20 to 100	12.3	19
2 to 20	13.8	21
0.4 to 2	14.0	25
less than 0.4	less than 0.06 mg/kg	less than 0.15 mg/kg

acceptability of results (95 % probability).

NOTE 2—The precision limits given in Table 2 are based on data published as Appendix III to Report of Committee D16, ASTM *Proceedings*, Am. Soc. Testing Mats., Vol 59, 1959, p. 514.

14.2 Repeatability is based on test results obtained by repetitive testing of a homogeneous sample by a single operator. Reproducibility is based on test results obtained by repetitive testing of different samples in different laboratories and different operators.

14.3 *Bias*—The bias of this test method has not been determined.

15. Keywords

15.1 benzene; spectrophotometry; thiophene

SUMMARY OF CHANGES

Committee D16 has identified the location of selected changes to this standard since the last date of issue that may impact the use of this standard.

- (1) Terminology D 4790 was added to Referenced Documents section. (2) Added section on Terminology.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).