



Designation: D 5198 – 92 (Reapproved 1997)

Standard Practice for Nitric Acid Digestion of Solid Waste¹

This standard is issued under the fixed designation D 5198; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This practice describes the digestion of solid waste using nitric acid for the subsequent determination of inorganic constituents by argon plasma emission spectroscopy or atomic absorption spectroscopy.

1.2 The following elements may be solubilized by this practice:

aluminum	manganese
beryllium	mercury
cadmium	nickel
chromium	phosphorus
copper	vanadium
iron	zinc
lead	

1.3 This practice is to be used when the concentrations of total recoverable elements are to be determined from a waste sample. Total recoverable elements may or may not be equivalent to total elements, depending on the element sought and the sample matrix. Recovery from refractory sample matrices, such as soils, is usually significantly less than total concentrations of the elements present.

NOTE 1—This practice has been used successfully for oily sludges and a municipal digested sludge standard [Environmental Protection Agency (EPA) Sample No. 397]. The practice may be applicable to some elements not listed above, such as arsenic, barium, selenium, cobalt, magnesium, and calcium. Refractory elements such as silicon, silver, and titanium are not solubilized by this practice.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For specific hazard statements, see Section 7.

2. Referenced Documents

2.1 ASTM Standards:

- D 1193 Specification for Reagent Water²
- D 2777 Practice for Determination of Precision and Bias of Applicable Methods of Committee D-19 on Water²
- D 3223 Test Method for Total Mercury in Water²
- D 3682 Test Method for Major and Minor Elements in Coal

¹ This practice is under the jurisdiction of ASTM Committee D34 on Waste Management and is the direct responsibility of Subcommittee D34.01.06 on Analytical Methods.

Current edition approved Jan. 9, 1992. Published March 1992.

² *Annual Book of ASTM Standards*, Vol 11.01.

- and Coke Ash by the Atomic Absorption Method³
- D 3919 Practice for Measuring Trace Elements in Water by Graphite Furnace Atomic Absorption Spectrophotometry²
- D 4503 Practice for the Dissolution of Solid Waste by Lithium Metaborate Fusion⁴
- D 4698 Practice for Total Digestion of Sediment Samples for Chemical Analysis of Various Metals⁵
- E 50 Practices for Apparatus, Reagents, and Safety Precautions for Chemical Analysis of Metals⁶

3. Summary of Practice

3.1 A weighed portion of the waste sample is mixed with 1 + 1 nitric acid (HNO₃) in an Erlenmeyer flask. The flask is heated for 2 h at 90 to 95°C to dissolve the elements of interest. After cooling, the contents of the flask are diluted with reagent water and filtered, and the filtrate is made up to appropriate volume for subsequent analysis.

4. Significance and Use

4.1 A knowledge of the inorganic composition of a waste is often required for the selection of appropriate waste disposal practices. Solid waste may exist in a variety of forms and contain a range of organic and inorganic constituents. This practice describes a digestion procedure which dissolves many of the toxic inorganic constituents and produces a solution suitable for determination by such techniques as atomic absorption spectroscopy, atomic emission spectroscopy, and so forth. The relatively large sample size aids representative sampling of heterogeneous wastes. The relatively small dilution factor allows lower detection limits than most other sample digestion methods. Volatile metals, such as lead and mercury, are not lost during this digestion procedure. Hydride-forming elements, such as arsenic and selenium, may be partially lost. Samples with total metal contents greater than 5 % may give low results. The analyst is responsible for determining whether this practice is applicable to the solid waste being tested.

5. Apparatus

- 5.1 *Analytical Balance*, capable of weighing to 0.01 g.
- 5.2 *Erlenmeyer Flasks*, 125 mL.
- 5.3 *Graduated Cylinder*, 50 mL.

³ *Annual Book of ASTM Standards*, Vol 05.05.

⁴ *Annual Book of ASTM Standards*, Vol 11.04.

⁵ *Annual Book of ASTM Standards*, Vol 11.02.

⁶ *Annual Book of ASTM Standards*, Vol 03.05.

 **D 5198**

5.4 *Electric Hot Plate*, adjustable, capable of maintaining a temperature of 90 to 95°C.

5.5 *Watch Glasses*.

5.6 *Thermometer*.

5.7 *Funnels*, glass or plastic.

5.8 *Volumetric Flasks*, glass-stoppered, 200 mL.

5.9 *Filter Paper*, quantitative, medium flow rate, Whatman No. 40 or equivalent.

6. Reagents

6.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.⁷ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

6.2 *Purity of Water*—Unless otherwise indicated, reference to water shall be understood to mean reagent water as defined by Type II of Specification D 1193.

6.3 *Nitric Acid*, concentrated, reagent grade.

6.4 *Nitric Acid (1+1)*—Add slowly, with stirring, 200 mL of concentrated nitric acid (HNO₃, sp gr 1.42) to 200 mL water. Cool the mixture and store in a clean pint glass bottle.

7. Hazards

7.1 Add the nitric acid mixture slowly, with swirling, to the sample. Samples containing carbonates may foam excessively during acid addition and result in loss of sample. Nitric acid may react violently with some samples containing organic material.

7.2 Addition of acid and sample digestion must be conducted in a hood with adequate ventilation and shielding to

⁷ “Reagent Chemicals, American Chemical Society Specifications,” Am. Chemical Soc., Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see “Reagent Chemicals and Standards,” by Joseph Rosin, D. Van Nostrand Co., Inc., New York, NY, and the “United States Pharmacopeia.”

avoid contact with nitrogen oxides, acid fumes, or toxic gases.

8. Procedure

8.1 Weigh 5 g of a thoroughly mixed waste sample to the nearest 0.01 g into a tared Erlenmeyer flask.

8.2 With a graduated cylinder, slowly add 25 mL of 1+1 nitric acid to the flask. Swirl the flask to wet the sample completely.

8.3 Carry a blank of 25 mL of 1+1 nitric acid through the procedure.

8.4 Place the flask on a cold hot plate, cover with a watch glass, and set the hot plate to maintain a temperature of 90 to 95°C.

8.5 Heat the flask and contents for 2 h, occasionally swirling the flask to wash down any sample adhering to the walls. Check the solution temperature with the thermometer and adjust the heat if necessary.

8.6 After 2 h, remove the flask from the hot plate and cool to room temperature. Add 50 mL of reagent water to the flask, washing down the flask walls during addition. Swirl the flask to mix the contents.

8.7 Filter the contents of the flask into a 200 mL volumetric flask. Rinse the flask and filter paper with several small portions of reagent water and add the rinsings to the volumetric flask.

8.8 Dilute the solution in the volumetric flask to the mark with reagent water and mix thoroughly. The solution is now ready for analysis.

9. Precision and Bias

9.1 No statement is made about either the precision or bias since this practice does not produce a test result. Appendix X1 contains representative results obtained with this practice and subsequent analysis.

10. Keywords

10.1 digestion; nitric acid; waste

APPENDIX

(Nonmandatory Information)

X1. REPRESENTATIVE ANALYSES OF SAMPLES AFTER NITRIC ACID DIGESTION



D 5198

TABLE X1.1 Nitric Acid Digestion—Analysis of Standard Sludge Sample (EPA Sample No. 397)

Element	Concentration, mg/kg		Percent Recovery
	Actual	Found	
Zn	1323	1300	98
Mn	205	235	115
Pb	519	500	96
Cd	20.8	24.1	116
Cr	204	218	107
Fe	16 155	16 400	102
V	13.0	12.7	98
Cu	1095	1130	103
Ni	198	186	94
Al	4558	4500	99
Ti	2121	95	4.5
P	11 573	11 800	102
Be	0.28	0.5	179
As	17	9	53
Ag	81	7.4	9.1
Hg	16.3	13.6	83

TABLE X1.3 Nitric Acid Digestion—Recovery of Used Motor Oil Matrix Spikes

Element	mg/kg		Percent Recovery
	Added	Found	
Mn	3.2	3.3	103
Pb	3.4	3.0	88
Cd	2.0	1.6	80
Cr	2.0	2.4	120
V	2.6	1.0	38
Cu	2.2	4.2	191
Ni	2.8	2.75	98
Ba	2.3	2.1	91
Be	1.6	1.3	81
Hg	0.08	0.09	113

TABLE X1.2 Nitric Acid Digestion—Replicate Analysis of Used Motor Oil, mg/kg

Element	Run 1	Run 2	Run 3
Zn	1170	1150	1170
Mn	2.5	2.1	2.1
Pb	43	40	35
Cd	0.6	0.5	0.4
Cr	2.8	2.1	2.3
Fe	74	66	66
V	1.32	1.04	0.92
Cu	3.8	3.6	3.6
Ni	0.7 ND ^A	0.7 ND	0.7 ND
Ba	0.6	0.5	0.5
Be	0.04 ND	0.04 ND	0.04 ND
Hg	0.08 ND	0.08 ND	0.08 ND

^A ND = not detected at the detection limit shown.

The American Society for Testing and Materials takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).