



Standard Test Method for Determining Plutonium by Controlled-Potential Coulometry in H₂SO₄ at a Platinum Working Electrode¹

This standard is issued under the fixed designation C 1165; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

^{ε1} NOTE—Section 2 and footnote 6 were editorially updated in July 2000.

1. Scope

1.1 This test method describes the determination of milligram quantities of plutonium in unirradiated uranium-plutonium mixed oxide having a U/Pu ratio range of 0.1 to 10. This test method is also applicable to plutonium metal, plutonium oxide, uranium-plutonium mixed carbide, various plutonium compounds including fluoride and chloride salts, and plutonium solutions.

1.2 The recommended amount of plutonium for each aliquant in the coulometric analysis is 5 to 10 mg. Precision worsens for lower amounts of plutonium, and elapsed time of electrolysis becomes impractical for higher amounts of plutonium.

1.3 The values stated in SI units are to be regarded as standard.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use. Specific precautionary statements are given in Section 8.*

2. Referenced Documents

2.1 ASTM Standards:

- C 757 Specification for Nuclear-Grade Plutonium Dioxide Powder, Sinterable²
- C 758 Test Methods for Chemical, Mass Spectrometric, Spectrochemical, Nuclear, and Radiochemical Analysis of Nuclear-Grade Plutonium Metal²
- C 759 Test Methods for Chemical, Mass Spectrometric, Spectrochemical, Nuclear, and Radiochemical Analysis of Nuclear-Grade Plutonium Nitrate Solutions²
- C 833 Specification for Sintered (Uranium-Plutonium) Dioxide Pellets²
- C 859 Terminology Relating to Nuclear Materials²
- C 1009 Guide for Establishing a Quality Assurance Pro-

gram for Analytical Chemistry Laboratories Within the Nuclear Industry²

C 1068 Guide for Qualification of Measurement Methods by a Laboratory Within the Nuclear Industry²

C 1108 Test Method for Plutonium by Controlled-Potential Coulometry²

C 1128 Guide for Preparation of Working Reference Materials for Use in the Analysis of Nuclear Fuel Cycle Materials²

C 1156 Guide for Establishing Calibration for a Measurement Method Used to Analyze Nuclear Fuel Cycle Materials²

C 1168 Practice for Preparation and Dissolution of Plutonium Materials for Analysis²

C 1210 Guide for Establishing a Measurement System Quality Control Program for Analytical Chemistry Laboratories Within the Nuclear Industry²

C 1297 Guide for Qualification of Laboratory Analysts for the Analysis of Nuclear Fuel Cycle Materials²

3. Summary of Test Method

3.1 In controlled-potential coulometry, the analyte reacts at an electrode having a maintained potential that precludes reactions of as many impurity components as is feasible. In the electrolysis, current decreases exponentially as the reaction proceeds until a selected background current is reached. The quantity of analyte reacted is calculable by Faraday's law. Detailed discussions of the theory and applications of this technique are presented in Refs (1)³ and (2).

3.2 Plutonium and many impurity element ions are initially reduced in a 0.5 M H₂SO₄ electrolyte at a platinum working electrode (3) maintained at +0.310 V versus a saturated calomel electrode (SCE). Plutonium is then oxidized to Pu(IV) at a potential of +0.670 V. The quantity of plutonium is calculated from the number of coulombs required for oxidation according to Faraday's law.

$$Q = \int_0^t i dt = nwF/M \tag{1}$$

Rearrangement to solve for w gives:

$$w = MQ/nF \tag{2}$$

¹ This test method is under the jurisdiction of ASTM Committee C-26 on Nuclear Fuel Cycle and is the direct responsibility of Subcommittee C26.05 on Methods of Test.

Current edition approved Nov. 30, 1990. Published July 1991.

² Annual Book of ASTM Standards, Vol 12.01.

³ The boldface numbers in parentheses refer to a list of references at the end of the text.

where:

- w = weight of Pu(III) oxidized to Pu(IV), g,
- M = gram-molecular mass of plutonium (adjusted for isotopic composition), grams/equivalent,
- Q = number of coulombs to oxidize Pu(III) to Pu(IV), coulombs,
- n = number of electron change to oxidize Pu(III) to Pu(IV) = 1, and
- F = Faraday constant, coulomb/equivalent.

3.3 An electrolyte of sulfuric acid, that selectively complexes Pu(IV), provides very reproducible electrolysis of Pu(III) to Pu(IV). In a 0.5 M H₂SO₄ electrolyte, the reduction potential of +0.310 V for conversion of Pu(IV and VI) to Pu(III) and the oxidation potential of +0.670 V for conversion of Pu(III) to Pu(IV) accounts for about 99.9 % (as calculated from the Nernst equation) conversion of the total plutonium in solution. There are few interferences at the selected potentials of the metallic impurities usually listed in specifications for fast breeder reactor (FBR) mixed oxide fuel. A chemical calibration of the coulometric system using the selected potentials technique is necessary to correct for the less than 100 % conversions of Pu(III) and Pu(IV).

3.4 Sulfuric acid is a convenient electrolyte since it is used for preliminary fuming of samples to volatilize interfering components (see 5.3 and 5.4). The preliminary fuming with sulfuric acid also serves to depolymerize any polymeric plutonium species, which tend to be electrolytically inactive (3).

4. Significance and Use

4.1 This test method is to be used to ascertain whether or not materials meet specifications for plutonium content or plutonium assay, or both.

4.2 A chemical calibration of the coulometer is necessary for accurate results.

5. Interferences

5.1 Categories of interferences are diverse metal ions that oxidize or reduce at the potential of +0.670 V used for the oxidation of Pu(III) to Pu(IV), organic matter, anions that complex plutonium, and oxygen.

5.2 The major interfering metallic impurity element, of those usually included in specifications for FBR mixed oxide

fuel, is iron (4). In the 0.5 M H₂SO₄ electrolyte, the Fe(II) – Fe(III) and Pu(III) – Pu(IV) couples have essentially the same E° value of +0.490 V. The iron interference, therefore, is quantitative and is corrected based on its measured value that can be determined by a spectrophotometric method (5). Alternatively, other techniques such as ICP, DCP or emission spectrometry can also be used if the iron content is sufficiently low. When the iron result is <20μ g/g, the lower limit of the spectrophotometric method, no correction is necessary. The best available method for iron determination is recommended since the uncertainty in the iron correction contributes to the uncertainty in the plutonium value.

5.3 Organic matter usually is not present in calcined mixed oxide fuel pellets nor in mixed oxide powder blends prepared using calcined uranium oxide and calcined plutonium oxide. However, it may be introduced as an impurity in reagents. The sulfuric acid fuming of reference material and of samples that precedes the coulometric analysis volatilizes most organic components.

5.4 The sulfuric acid fuming volatilizes nitrate, nitrite, fluoride, and chloride, that are introduced by the use of a nitric-hydrofluoric acid mixture or acid mixtures containing chloride for the dissolution of samples and interfere in the coulometric determination of plutonium.

5.5 Oxygen interferes and must be purged continuously from both the solution and atmosphere in the electrolysis cell with an oxygen-free inert gas before and during the electrolysis.

NOTE 1—The purge gas tube extends through the cell cover and is positioned approximately 1 cm above the sample solution in the cell. The inert gas flow is maintained at a flow rate that causes a dimple to be seen on the surface of the solution with the stirrer off. The inert gas flow rate should be such that no splashing occurs.

6. Apparatus

6.1 *Controlled-Potential Coulometer*—A potentiostat having stable potential control at approximately 200 mA and 20 V and an integrator capable of 0.05 % reproducibility are required. The linearity of the integrator should be better than 0.1 % for the selected range.

NOTE 2—To obtain maximum precision, it is recommended that the reference and sample aliquants contain approximately the same amount of plutonium.⁴

6.2 *Cell Assembly*—A cell assembly similar to the one described in Ref (5) has been used satisfactorily. Cell design is very critical in controlled-potential coulometry. There are many factors that must be considered in choosing or designing a cell assembly. It is beyond the scope of this test method to describe all of the factors that should be considered. A thorough detailed discussion of electrolysis cell design is presented in Ref (2).

NOTE 3—Drawing (see Fig. 1) of a cell design that has been successfully used at the Los Alamos National Laboratory. The titration cell consists of a 50 mL cut off beaker.

6.3 Timer or stopwatch for measuring electrolysis times




FIG. 1 Example of a Cell Design Used at Los Alamos National Laboratory (LANL)

⁴ Apparatus manufactured by the EG&G Princeton Applied Research Corp., has been found satisfactory for this purpose.

(capable of measuring in seconds).

7. Reagents

7.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society where such specifications are available.⁵ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

7.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean distilled or deionized water.

7.3 *Argon, Oxygen-Free (99.99 %)*—Helium, nitrogen, or other pure inert gas may be used.

7.4 *Hydrochloric Acid (HCl, 6 M)*—Add 500 mL of concentrated HCl (sp gr 1.19) to less than 500 mL of water and dilute to 1 L with water.

7.5 *Sulfuric Acid (sp gr 1.84)*—Concentrated H₂SO₄ (sp gr 1.84).

7.6 *Sulfuric Acid (3 M)*—Add 168 mL of concentrated H₂SO₄ (sp gr 1.84) to water, while stirring, and dilute to 1 L with water.

7.7 *Sulfuric Acid (0.5 M)*—Add 28 mL of concentrated H₂SO₄ (sp gr 1.84) to water, while stirring, and dilute to 1 L with water.

7.8 *Plutonium Reference Solution*—Dissolve a weighed quantity (balance capable of weighing to 0.01 mg) of 0.5 to 1 g of NBL CRM 126 metal (or its replacement) cleaned per certificate directions in 6 M HCl. Use a sufficient amount of 6 M HCl to maintain an acid concentration of 1 to 2 molar. Completely transfer the solution with 0.5 M H₂SO₄ rinses to a tared container, dilute to 100 to 200 g with 0.5 M H₂SO₄ (to give a plutonium concentration of 5 mg/g), and weigh.

NOTE 4—A tared polyethylene bottle has been used successfully to dispense weighed aliquants.

7.8.1 Dispense weighed 1 to 2 g aliquants, each containing accurately known 5 to 10 mg quantities of plutonium, to individual electrolysis cells or vials for subsequent use in chemical calibration.

7.8.2 Prior to using, add 0.5 mL of 3 M H₂SO₄ and fume to dryness.

7.8.3 After cooling, redissolve using a minimal amount of 0.5 M H₂SO₄ and again fume to dryness.

7.8.4 Repeat 7.8.3.

8. Safety Precautions

8.1 *Committee C-26 Safeguards Statement*⁶:

8.1.1 The materials (nuclear grade plutonium metal, plutonium oxide powder, plutonium nitrate solutions, and mixed

oxide and carbide powders and pellets) to which this test method applies, are subject to nuclear safeguards regulations governing their possession and use. This test method has been designated as technically acceptable for generating safeguards accountability measurement data.

8.1.2 When used in conjunction with appropriate certified reference materials (CRMs), this test method can demonstrate traceability to the national measurements base. However, adherence to this test method does not automatically guarantee regulatory acceptance of the resulting safeguards measurements. It remains the sole responsibility of the user of this test method to ensure that its application to safeguards has the approval of the proper regulatory authorities.

9. Preparation of Apparatus

9.1 Verify proper equipment operation by performing an electrical calibration according to manufacturers' specifications on each day that the instrument is used.

10. Calibration

10.1 If not done previously as recommended in 7.8.1, completely transfer one of the dispensed aliquants, containing 5 to 10 mg of plutonium of the plutonium reference solution, to a cell using 0.5 M H₂SO₄ rinses and place platinum working electrode in the cell. Using 0.5 M H₂SO₄, completely immerse the working electrode. (See Note 8.)

10.2 Rinse the exterior surfaces of the counter and reference electrode salt bridges (for example, high-silica tubes) with 0.5 M H₂SO₄.

10.3 Raise the cell into position firmly against the cell cover to ensure a tight fit. Purge the cell atmosphere with flowing argon or other inert gas. (See Note 1.)

10.4 Immediately connect the cell electrodes to the coulometer; begin stirring.

10.5 Reduce Pu(IV) to Pu(III) at + 0.310 V until the current decreases to 30 μA.

10.6 Reset the integrator and start timer.

10.7 Oxidize Pu(III) to Pu(IV) at + 0.670 V until the current decreases to 30 μA. Record the coulomb accumulation and elapsed time.

NOTE 5—All standards (reference material) and samples should be freshly fumed (within 4 h) prior to analysis.

10.8 Remove the solution and thoroughly rinse the cell and electrodes with 0.5 M H₂SO₄.

10.9 Repeat 10.1-10.8 to attain a desired precision level for the calibration.

NOTE 6—A recommended practice would be to intersperse standards (reference material) and samples during the time the analyses are being done.

10.10 Calculate the calibration factor F by

$$F = M/(C_C - C_B) \quad (3)$$

where:

F = calibration factor, milligrams plutonium per coulomb,

M = mass of plutonium in calibration reference aliquant, milligrams,

⁵ Reagent Chemicals, American Chemical Society Specifications, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see Analar Standards for Laboratory Chemicals, BDH Ltd., Poole, Dorset, U.K., and the United States Pharmacopeia and National Formulary, U.S. Pharmaceutical Convention, Inc. (USPC), Rockville, MD.

⁶ Based upon Committee C26 Safeguards Matrix (C 1009, C 1068, C 1128, C 1156, C 1210, and C 1297).

C_C = coulombs measured at 0.670 V electrolysis for calibration reference aliquant, and
 C_B = coulombs for blank measurement. Use the coulomb value obtained on the blank for the elapsed time (to the nearest minute) as that required for the reference aliquant oxidation time.

11. Procedure

11.1 Blanks:

11.1.1 Obtain reproducible blank measurements on each individual platinum electrode by following 11.1.2-11.1.8.

NOTE 7—Two platinum working electrodes are recommended to increase sample throughput by alternating the electrodes. While one electrode is being used in an electrolysis, the other electrode is being cleaned by sitting in a beaker of hot concentrated nitric acid. The electrode that is being cleaned is rinsed thoroughly with water and 0.5 M sulfuric acid prior to its use.

11.1.2 Add 0.5 M H_2SO_4 to the cell to completely immerse the working electrode.

NOTE 8—Avoid overfilling the cell. Fill only to the top of the platinum gauze working electrode. Overfilling the cell will result in longer electrolysis times and larger background currents.

11.1.3 Rinse the counter and reference electrode salt bridges (high-silica tubes) with 0.5 M H_2SO_4 .

11.1.4 Raise the cell into position firmly against the cell cover to ensure a tight fit. Purge the cell atmosphere with flowing argon or other inert gas. (See Note 1.)

11.1.5 Immediately connect the cell electrodes to the coulometer; begin stirring.

11.1.6 Electrolyze the blank at 0.310 V until a 30- μ A current is obtained.

11.1.7 Start the timer, and electrolyze the blank at 0.670 V for a period of time that is consistent with sample electrolysis times.

11.1.8 Record the number of coulombs at elapsed electrolysis times consistent with sample electrolysis times.

11.1.9 Following blanks, run a plutonium cell conditioner sample to equilibrate the cell prior to running standards (reference material) and samples.

NOTE 9—A plutonium cell conditioner sample is a plutonium solution that is run through the complete reduction/oxidation cycle but is not used for calculation purposes. Experience has shown that if a plutonium cell conditioner is not run, the initial plutonium result will be low. A possible cause for this effect is migration of plutonium into the high silica tubes until equilibration is attained.

11.2 Sample Analysis:

11.2.1 The plutonium-containing material may be dissolved using the appropriate dissolution procedure described in Practice C 1168.

11.2.2 After transferring and diluting, weigh aliquants containing 5 to 10 mg of plutonium.

11.2.3 Add 0.5 mL of 3 M H_2SO_4 to each aliquant and fume to dryness.

11.2.4 After cooling, dissolve the sample using a minimal amount of 0.5 M H_2SO_4 and again fume to dryness.

11.2.5 Repeat 11.2.4.

11.2.6 Dissolve the sample using a minimal amount of 0.5 M H_2SO_4 .

11.2.7 Place a platinum working electrode in the cell and completely immerse the working electrode using 0.5 M H_2SO_4 .

11.2.8 Proceed with the coulometric analysis of one or more aliquants by following 10.2-10.8.

11.2.9 Correct for the iron content of the sample, which has been determined using the recommended spectrophotometric procedure or a suitable alternate procedure.

12. Calculation of Sample Result

12.1 Calculate the plutonium content of the sample by

$$Pu = (D) (A_S/A_R) F(C_S - C_B)/M_S \quad (4)$$

where:

Pu = result, gram plutonium per gram sample,

D = dilution factor, grams of diluted sample/grams of aliquant analyzed,

A_S = atomic weight of plutonium in sample,

A_R = atomic weight of plutonium in plutonium metal reference material,

F = average calibration factor, milligram plutonium per coulomb (see 10.10),

C_S = coulombs measured for 0.670 V electrolysis for sample aliquant,

C_B = coulombs for blank measurement (same elapsed time to the nearest minute as for sample), and

M_S = mass of solid sample initially dissolved, milligrams.

12.2 Calculate the correction for iron by

$$Fe_c = (10^{-6})(Fe)(A_S/55.85) \quad (5)$$

where:

Fe_c = correction for iron, gram plutonium/gram of sample,

Fe = micrograms iron/gram of sample, and

A_S = atomic weight of plutonium in sample.

12.3 Calculate the corrected plutonium content, Pu_c , of the sample by

$$Pu_c = Pu - Fe_c \quad (6)$$

13. Precision and Bias

13.1 For a single measurement on an aliquant, the estimated repeatability relative standard deviation is 0.10 % and the estimated reproducibility relative standard deviation is 0.15 %. These estimates are based on the analysis of five samples, four aliquants each, by each of six laboratories (6) and the analysis of 153 aliquants involving nine distinct dissolutions of a control sample at one laboratory. If more than one aliquant is measured (see 11.2.8) and the average reported, the repeatability and reproducibility relative standard deviations are 0.10/ \sqrt{n} % and 0.15/ \sqrt{n} % respectively, where n is the number of measurements in the average.

13.2 Comparison with a potentiometric method, a photometric method, and with 100 % impurities data indicate that the coulometric method is essentially unbiased.

14. Keywords

14.1 controlled-potential coulometry; plutonium analysis; plutonium at platinum electrode; plutonium in sulphuric acid; plutonium-uranium mixtures

REFERENCES

- (1) Shults, W. D., "Coulometric Methods", Standard Methods of Chemical Analysis, F. J. Welcher, ed., D. Van Nostrand Co., Inc., Princeton, NJ, Chapter 23, Vol. IIIA, 1967.
- (2) Harrar, J. E., "Techniques, Apparatus, and Analytical Applications of Controlled-Potential Coulometry," Electroanalytical Chemistry, Vol. 8, A. J. Bard, ed., Marcel Dekker, New York, NY, Chapter 1, 1975.
- (3) Shults, W. D., "Applications of Controlled-Potential Coulometry to the Determination of Plutonium—A Review," Talanta, Vol. 10, 1963, p. 833.
- (4) Stokely, J. R., Jr., and Shults, W. D., "Controlled Potential Coulometric Determination of Plutonium in the Presence of Iron," Analytical Chemistry, Vol. 43, 1971, pp. 603–605.
- (5) Rein, J. E., Matlack, G. M., Waterbury, G. R., Phelps, R. T., and Metz, C. F., eds., "Methods of Chemical Analysis for FBR Uranium-Plutonium Mixed Oxide Fuel and Source Materials," Los Alamos Scientific Laboratory Report LA-4622, 1971.
- (6) Rein, J. E., Ziegler, R. K., and Metz, C. F., "LMFBR/FFTF Fuel Development Analytical Chemistry Program (Phase II)," USAEC Report LA-4407, 1970.

The American Society for Testing and Materials takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).