



Standard Guide for Selection and Use of Neutron Sensors for Determining Neutron Spectra Employed in Radiation-Hardness Testing of Electronics¹

This standard is issued under the fixed designation E 720; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

This standard has been approved for use by agencies of the Department of Defense.

1. Scope

1.1 This guide covers the selection and use of neutron-activation detector materials to be employed in neutron spectra adjustment techniques used for radiation-hardness testing of electronic semiconductor devices. Sensors are described that have been used at many radiation hardness-testing facilities, and comments are offered in table footnotes concerning the appropriateness of each reaction as judged by its cross-section accuracy, ease of use as a sensor, and by past successful application. This guide also discusses the fluence-uniformity, neutron self-shielding, and fluence-depression corrections that need to be considered in choosing the sensor thickness, the sensor covers, and the sensor locations. These considerations are relevant for the determination of neutron spectra from assemblies such as TRIGA- and Godiva-type reactors and from Californium irradiators. This guide may also be applicable to other broad energy distribution sources up to 20 MeV.

NOTE 1—For definitions on terminology used in this guide, see Terminology E 170.

1.2 This guide also covers the measurement of the gamma-ray or beta-ray emission rates from the activation foils and other sensors as well as the calculation of the absolute specific activities of these foils. The principal measurement technique is high-resolution gamma-ray spectrometry. The activities are used in the determination of the energy-fluence spectrum of the neutron source. See Guide E 721.

1.3 Details of measurement and analysis are covered as follows:

1.3.1 Corrections involved in measuring the sensor activities include those for finite sensor size and thickness in the calibration of the gamma-ray detector, for pulse-height analyzer deadtime and pulse-pileup losses, and for background radioactivity.

1.3.2 The primary method for detector calibration that uses secondary standard gamma-ray emitting sources is considered in this guide and in General Methods E 181. In addition, an alternative method in which the sensors are activated in the known spectrum of a benchmark neutron field is discussed in Guide E 1018.

1.3.3 A data analysis method is presented which accounts for the following: detector efficiency; background subtraction; irradiation, waiting, and counting times; fission yields and gamma-ray branching ratios; and self-absorption of gamma rays and neutrons in the sensors.

1.4 The values stated in SI units are to be regarded as the standard.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 General considerations of neutron-activation detectors discussed in Practice E 261, Test Method E 262, and Guides E 721 and E 844 are applicable to this guide. Background information for applying this guide are given in these and other relevant standards as follows:

2.2 ASTM Standards:

E 170 Terminology Relating to Radiation Measurements and Dosimetry²

E 181 Test Methods for Detector Calibration and Analysis of Radionuclides²

E 261 Practice for Determining Neutron Fluence Rate, Fluence, and Spectra by Radioactivation Techniques²

E 262 Test Method for Determining Thermal Neutron Reaction and Fluence Rates by Radioactivation Techniques²

E 263 Test Method for Measuring Fast-Neutron Reaction Rates by Radioactivation of Iron²

E 264 Test Method for Measuring Fast-Neutron Reaction Rates by Radioactivation of Nickel²

¹ This guide is under the jurisdiction of ASTM Committee E10 on Nuclear Technology and Applications and is the direct responsibility of Subcommittee E10.07 on Radiation Effects on Electronic Materials and Devices and Pulsed Radiation Effects.

Current edition approved June 10, 2002. Published September 2002. Originally published as E 720 – 80. Last previous edition E 720 – 94.

² *Annual Book of ASTM Standards*, Vol 12.02.

- E 265 Test Method for Measuring Reaction Rates and Fast-Neutron Fluences by Radioactivation of Sulfur-32²
- E 266 Test Method for Measuring Fast-Neutron Reaction Rates by Radioactivation of Aluminum²
- E 343 Test Method for Measuring Reaction Rates by Analysis of Molybdenum-99 Radioactivity from Fission Dosimeters²
- E 393 Test Method for Measuring Reaction Rates by Analysis of Barium-140 from Fission Dosimeters²
- E 496 Test Method for Measuring Neutron Fluence Rate and Average Energy from ³H(*d,n*) ⁴He Neutron Generators by Radioactivation Techniques²
- E 704 Test Method for Measuring Reaction Rates by Radioactivation of Uranium-238²
- E 705 Test Method for Measuring Reaction Rates by Radioactivation of Neptunium-237²
- E 721 Guide for Determining Neutron Energy Spectra from Neutron Sensors for Radiation-Hardness Testing of Electronics²
- E 844 Guide for Sensor Set Design and Irradiation for Reactor Surveillance, E 706 (IIC)²
- E 944 Guide for Application of Neutron Spectrum Adjustment Methods in Reactor Surveillance, E706 (IIA)²
- E 1018 Guide for Application of ASTM Evaluated Cross Section Data File, E 706 (IIB)²
- E 1297 Test Method for Measuring Fast-Neutron Reaction Rates by Radioactivation of Niobium²

3. Significance and Use

3.1 Because of the wide variety of materials being used in neutron-activation measurements, this guide is presented with the objective of bringing improved uniformity to the specific field of interest here: hardness testing of electronics primarily in critical assembly reactor environments.

NOTE 2—Some of the techniques discussed are useful for 14-MeV dosimetry. See Test Method E 496 for activation detector materials suitable for 14-MeV neutron effects testing.

NOTE 3—The materials recommended in this guide are suitable for ²⁵²Cf or other weak source effects testing provided the fluence is sufficient to generate countable activities.

3.2 This guide is organized into two overlapping subjects; the criteria used for sensor selection, and the procedures used to ensure the proper determination of activities for determination of neutron spectra. See Terminology E 170 and General Methods E 181. Determination of neutron spectra with activation sensor data is discussed in Guides E 721 and E 944.

4. Foil Sets

4.1 Reactions Considered:

4.1.1 Neutron-induced reactions appropriate for this guide are listed in Table 1. The table includes most of the reactions used in this field. Those not marked with an asterisk are recommended because of their demonstrated compatibility with other reactions used in spectrum adjustment determinations. This compatibility is primarily based on experience with the ENDF/B-VI (1), and IRDF-90 (2) cross-sections. These recommendations may change modestly as revisions are made in the ENDF/B-VI and IRDF-90 dosimetry cross sections. Other reactions may be useful in particular circumstances with

appropriate care. It is important that the user take full account of both the footnotes attached to each reaction and the discussions in the body of the text about individual reactions when implementing the foil-activation technique.

4.1.2 The four paired columns under the labels fast burst (3) and “TRIGA (4) Type” list the energy ranges within which 95 % of the response occurs for these two representative spectra. These limits are just a guide because the response often varies widely within each range. The response limits for an idealized fission spectrum with no 1/*E* tail can be much different (shifted toward higher energy) for resonance reactions. For example, in a Watt fission spectrum the ¹⁹⁷Au(*n,γ*)¹⁹⁸Au has a 95 % response between 5.0×10^{-2} and 2.7 MeV. The recommended foil mass column gives values that are designed to minimize self-absorption, self-shielding, and other corrections, provided the foils are 1.27 cm in diameter. The $E_f \approx 0$ fission foils, ²³⁵U and ²³⁹Pu, have similar cross-section shapes. However, the ²³⁵U foil is preferred since it is less expensive and is much less of a health hazard than ²³⁹Pu. In addition, when measuring soft (TRIGA) spectra, the ²³⁵U foil is useful in determining the correction for the ²³⁵U impurity in the ²³⁸U foil (which is readily available with about 400 ppm or less ²³⁵U impurity).

4.1.3 Although sulfur is listed and is used widely as a monitor foil, it is the only recommended sensor requiring beta particle detection and, therefore, requires a different calibration and counting technique. The ⁵⁸Ni(*n,p*)⁵⁸Co reaction has about the same threshold energy and, therefore, can be used instead of the ³²S(*n,p*)³²P if it acquires sufficient activity. Many facilities use sulfur as a routine monitor because its two-week half-life allows a convenient period for counting and permits reuse of the sensor after 6 to 9 months. Automated beta counters are commercially available. Neither nickel nor sulfur should be counted for the (*n,p*) reaction products immediately after irradiation because for nickel the ⁵⁸Co must build up through a metastable state, and for sulfur there are competing reactions. According to Test Method E 264 the waiting period for ⁵⁸Co should be 4 days. For ³²P, Test Method E 265 recommends waiting 24 h. Corrections can be made for shorter waiting periods.

4.1.4 In selecting dosimetry reactions one should consider the validation of the cross sections and associated uncertainty as demonstrated in the ²³⁵U thermal fission and the ²⁵²Cf spontaneous fission benchmark neutron fields. 20 provides a recent comparison of the measured and calculated spectrum-averaged cross sections for these benchmark fields.

4.1.5 Some frequently used reactions have shown relatively consistent deviations of measured to calculated activity ratios in many different spectra determinations. For example, when ENDF/B-V cross sections are used in the reaction ⁶³Cu(*n,γ*)⁶⁴Cu, the calculated activity is usually low, and an adjustment code will try to raise the spectrum in the vicinity of Cu resonances. In fact, however, this consistent behavior indicates that the tabulated cross-section values in some important energy region are too small. The analyst must then choose one of the following alternatives: (1) leave out reactions which have demonstrated consistent deviations; (2) seek better cross-section sets; (3) assign wide error bars or low statistical

TABLE 1 Activation Foils

Reaction	Fast Burst ^A		TRIGA Type ^A		E _γ , keV	Gamma/Reaction ^B (Fast Fission Yield,% ^C)	T _{1/2} ^B	Recommended Foil Mass, g ^D	Footnotes
	E _L , MeV	E _H , MeV	E _L , MeV	E _H , MeV					
¹⁹⁷ Au(n,γ) ¹⁹⁸ Au	4.00 – 6	7.20 – 4	3.80 – 6	9.20 – 6	411.8	0.956	2.694 days	0.06	E,F,G
⁵⁹ Co(n,γ) ⁶⁰ Co	7.60 – 6	4.50 – 4	6.90 – 7	1.43 – 4	1173.2	0.9998	5.271 years	0.06	E,G
⁵⁸ Fe(n,γ) ⁵⁹ Fe	1.00 – 6	2.10 + 0	5.25 – 7	1.00 – 2	1332.5	0.9998	44.5 days	0.15	E,H
					1099.2	0.565			
⁵⁵ Mn(n,γ) ⁵⁶ Mn	5.25 – 7	6.60 – 1	4.75 – 7	1.10 – 3	1291.6	0.432	2.58 h	0.05	E,F,I
					846.8	0.989			
					1810.7	0.272			
⁶³ Cu(n,γ) ⁶⁴ Cu	1.15 – 6	2.30 + 0	5.25 – 7	9.60 – 3	1345.9	0.0049	12.7 h	0.15	E
²³ Na(n,γ) ²⁴ Na	6.30 – 7	2.00 + 0	5.25 – 7	3.00 – 3	1368.6	1.00	14.96 h	0.10	E,J,K
⁴⁵ Sc(n,γ) ⁴⁶ Sc	4.25 – 7	1.00 + 0	4.00 – 7	4.75 – 4	1120.5	1.00	83.81 days	0.05	E
²³⁵ U(n,γ) ²³⁶ U	9.20 – 2	4.70 + 0	6.30 – 4	3.80 + 0	1596.2	0.954 (6.105)	40.27 h	0.30	E,L,M
²³⁵ U(n,γ) ²³⁶ U	9.20 – 2	4.70 + 0	6.30 – 4	3.80 + 0	724.2	0.441 (6.363)	64.02 days	0.60	E,M
²³⁹ Pu(n,γ) ²⁴⁰ Pu	1.43 – 1	4.80 + 0	8.80 – 4	4.30 + 0	1596.2	0.954 (5.326)	40.27 h	1.00	E,L,M
²³⁹ Pu(n,γ) ²⁴⁰ Pu	1.43 – 1	4.80 + 0	8.80 – 4	4.30 + 0	724.2	0.441 (4.685)	64.02 days	0.60	E,M
⁹³ Nb(n,n') ^{93m} Nb	8.40 – 1	5.70 + 0	1.00 + 0	5.50 + 0	16.6	0.115	16.13 years		N
¹⁰³ Rh(n,n') ^{103m} Rh	5.50 – 1	5.70 + 0	6.90 – 1	5.70 + 0	39.8	0.068	56.1 min		N
²³⁷ Np(n,γ) ²³⁸ Np	5.75 – 1	5.60 + 0	6.60 – 1	5.50 + 0	1596.2	0.954 (5.489)	40.27 h	0.60	E,L,M,O
²³⁷ Np(n,γ) ²³⁸ Np	5.75 – 1	5.60 + 0	6.60 – 1	5.50 + 0	724.2	0.441 (5.699)	64.02 days	0.60	E,M
¹¹⁵ In(n,n') ^{115m} In	1.00 + 0	6.00 + 0	1.20 + 0	5.80 + 0	336.2	0.459	4.49 h	0.12	
²³⁸ U(n,γ) ²³⁹ U	1.50 + 0	6.90 + 0	1.50 + 0	6.60 + 0	1596.2	0.954 (5.948)	40.27 h	1.00	E,L,M,P
²³⁸ U(n,γ) ²³⁹ U	1.50 + 0	6.90 + 0	1.50 + 0	6.60 + 0	724.2	0.441 (5.105)	64.02 days	1.00	E,M
²³² Th(n,γ) ²³³ Th	1.50 + 0	7.40 + 0	1.50 + 0	7.10 + 0	537.3	0.244 (7.704)	12.75 days	1.00	E,L,Q
²³² Th(n,γ) ²³³ Th	1.50 + 0	7.40 + 0	1.50 + 0	7.10 + 0	724.2	0.441 (5.374)	64.02 days	1.00	E,M
⁵⁴ Fe(n,p) ⁵⁴ Mn	2.30 + 0	7.70 + 0	2.30 + 0	7.40 + 0	834.8	1.00	312.1 days	0.15	E
⁵⁸ Ni(n,p) ⁵⁸ Co	2.00 + 0	7.60 + 0	2.00 + 0	7.30 + 0	810.8	0.995	70.8 days	0.30	E
⁴⁷ Ti(n,p) ⁴⁷ Sc	1.90 + 0	7.60 + 0	1.90 + 0	7.30 + 0	159.4	0.683	3.35 days	0.15	E,R,S
³² S(n,p) ³² P	2.40 + 0	7.50 + 0	2.30 + 0	7.30 + 0	1710.(beta)	1.00 (beta)	14.28 days	...	T
⁶⁴ Zn(n,p) ⁶⁴ Cu	2.60 + 0	7.70 + 0	2.60 + 0	7.40 + 0	1345.9	0.0049	12.7 h	0.30	E
²⁷ Al(n,p) ²⁷ Mg	3.50 + 0	9.40 + 0	3.40 + 0	9.20 + 0	843.8	0.718	9.46 min	0.30	E
⁴⁶ Ti(n,p) ⁴⁶ Sc	3.80 + 0	9.60 + 0	3.70 + 0	9.20 + 0	1120.5	1.00	83.81 days	0.15	E,R
⁵⁶ Fe(n,p) ⁵⁶ Mn	5.50 + 0	1.14 + 1	5.50 + 0	1.10 + 1	846.8	0.989	2.58 h	0.15	E,U
²⁴ Mg(n,p) ²⁴ Na	6.50 + 0	1.17 + 1	6.50 + 0	1.13 + 1	1368.6	1.00	14.96 h	0.03	E,K
²⁷ Al(n,α) ²⁴ Na	6.50 + 0	1.21 + 1	6.50 + 0	1.17 + 1	1368.6	1.00	14.96 h	0.30	EK
⁴⁸ Ti(n,p) ⁴⁸ Sc	5.90 + 0	1.24 + 1	5.90 + 0	1.20 + 1	983.5	1.00	43.7 h	0.15	E
⁹³ Nb(n,2n) ^{92m} Nb	9.70 + 0	1.45 + 1	9.40 + 0	1.40 + 1	1037.5	0.975	10.15 days	0.25	E
					1312.1	1.00			
					934.4	0.992			
¹²⁷ I(n,2n) ¹²⁶ I	9.70 + 0	1.47 + 1	9.70 + 0	1.43 + 1	388.6	0.341	13.02 days		
⁶⁵ Cu(n,2n) ⁶⁴ Cu	1.08 + 1	1.57 + 1	1.07 + 1	1.53 + 1	666.	0.331	12.7 h	0.15	E,N
					1345.9	0.0049			
⁶³ Cu(n,2n) ⁶² Cu	1.19 + 1	1.66 + 1	1.19 + 1	1.63 + 1	875.7	0.00150	9.74 min	0.15	E,H
⁹⁰ Zr(n,2n) ⁸⁹ Zr	1.28 + 1	1.69 + 1	1.27 + 1	1.67 + 1	909.1	0.999	78.4 h	0.10	
⁵⁸ Ni(n,2n) ⁵⁷ Ni	1.32 + 1	1.71 + 1	1.31 + 1	1.69 + 1	1377.6	0.80	1.49 days	0.30	

^A Energy limits inside of which 95 % of the detector response occurs for each reaction (see Practice E 261 and Ref 6). The foils are assumed to have Cd covers as described in Footnote E.

^B Data from Ref for isotopes with atomic weight greater than or equal to 45, from Ref for isotopes with atomic weight less than 45.

^C Fission yields can be found in Ref 7. Because Ref 7 documentation is still in draft form, the fission yield data in the table is taken from Ref .

^D Choice of mass is based on assumed foil diameter of 1.27 cm.

^E Cd covers 0.5 to 1-mm thicknesses. Pairs of bare and Cd-covered foils are advantageous for resonance reactions.

^F Use ⁵⁹Co instead of ¹⁹⁷Au and ⁵⁵Mn for very long irradiations.

^G Use dilute aluminum-gold alloy (<0.2 % Au) when possible.

^H Do not count the 0.511 line.

^I Resonance structure differs in ENDF/B-VI from ENDF/B-V.

^J Use in the form of NaCl.

^K The 1986 edition of Ref has a typographical error for the half-life of ²⁴Na. The correct number can be found in previous editions. The correct number can also be found in Ref 10.

^L This is the 40.27-h daughter of 12.75-day ¹⁴⁰Ba. Wait 5 days for maximum decay rate (see Test Method E 393).

^M E_r = 0.01 MeV shielded with ¹⁰B sphere. (Use of ¹⁰B shield is important for soft (TRIGA) spectra where Φ(E < 0.01 MeV) will otherwise dominate).

^N Precautions must be taken in counting because of the low gamma-ray energy. See Test Method E 1297.

^O If a ¹⁰B sphere is used for the ²³⁹Pu foil, then a ¹⁰B sphere should also be used for the ²³⁷Np foil so that correction for ²³⁹Pu impurity in the ²³⁷Np foil can be made.

^P If a ¹⁰B sphere is used for the ²³⁵U foil, then a ¹⁰B sphere should also be used for the ²³⁸U foil so that correction for ²³⁵U impurity in the ²³⁸U foil can be made.

^Q Radioactivity of ²³²Th interferes with the ¹⁴⁰La line.

^R At high energies (>10 MeV), account for (n,np) contributions from higher atomic number Ti isotopes.

^S See Refs 8 and 9.

^T Requires β counting techniques, see Test Method E 265.

^U Maximum Mn impurity = 0.001 %, Cd covered. Do not use ⁵⁶Fe foil for long irradiations.

* Not recommended for use at this time either because of large uncertainties or because of conflicts with other reactions during spectrum adjustment procedures.

weight to these reactions. It is recommended that the first option be chosen because a sufficient number of well-established cross sections do exist to satisfactorily determine

fast reactor spectra. Furthermore, if the cross section for a particular reaction is not well established, and it is assigned too large a weight in the spectrum adjustment procedure, the final

spectrum can be severely distorted. Other suspect reactions are noted in Table 1 with an asterisk.

NOTE 4—Some of the reactions not recommended at this time (on the basis of inconsistencies among recommended cross sections) may be upgraded when more recent evaluations are applied to a wide range of neutron spectra.

4.2 Foil Impurities:

4.2.1 Foil impurities are especially serious for a moderated source (TRIGA reactor) when impurity leads to the same reaction product by way of thermal-neutron capture. Some examples of these foils, with impurities in parentheses, are ^{238}U (^{235}U), ^{27}Al (^{23}Na), ^{56}Fe (^{55}Mn), and ^{27}Mg (^{23}Na).

4.2.2 For a soft spectrum, such as the TRIGA J-tube spectrum [boral (boron-aluminum alloy) shielded], the number of fissions in the ^{235}U foil (Cd covered) is about 100 times the number occurring in the ^{238}U foil; therefore, the ^{238}U must have an impurity level of ^{235}U of no more than about 200 ppm for an uncertainty of 2 % or less in determining accurately the ^{238}U activity. Higher impurity levels of ^{235}U can be tolerated for Godiva-type reactors where the fluence below 10 keV is much lower, or with TRIGA-type reactors if the ^{235}U foil data are used for correcting the ^{238}U activity.

4.2.3 When the ^{56}Fe foil (Cd covered) is used in a TRIGA spectrum, it should have no more than 10 ppm ^{55}Mn impurity to keep the contribution from the $^{55}\text{Mn}(n,\gamma)^{56}\text{Mn}$ reaction to less than 2 %. Similarly, the ^{55}Mn impurity should be no more than 100 ppm when using the ^{56}Fe foil at 50 cm from a Godiva-type reactor (which is approximately 2 m above the concrete floor) in order to achieve the same level of accuracy. Data from a ^{55}Mn foil (Cd covered) can be used to correct the ^{56}Fe data if the impurity correction is ≤ 20 % of the total (n,p) activation, and the percent of manganese in iron is accurately known.

4.3 Influence of Nuclear Data on Foil Selection:

4.3.1 Since the total number of interactions is deduced from an absolute specific activity determination, that activity should be determined with good accuracy (of the order of 5 %), and the foils selected should have gamma-ray yields known to the same or better accuracy. Some of the factors involved in determining these yields include conversion-electron production, branching ratio to a given energy level, and fission yield.

4.3.2 The 1596-keV gamma-ray transition from ^{140}La produced by ^{232}Th fission is not usually useful because of interference from ^{232}Th radioactivity. This often has led to the use of the 537-keV transition from the ^{140}Ba precursor of ^{140}La , having a gamma-transition probability of 0.244 per ^{140}Ba decay. The use of ^{140}Ba generally requires the chemical separation of this isotope from the rest of the fission products so that the 537-keV line can be seen above competing lines. See Test Method E 393.

4.3.3 The choice of element, and hence the gamma-ray transition, directly influences the accuracy of determining the specific activity induced by neutron irradiation. It also influences the final choice of foil thickness, in that the selection of an element resulting in a low-energy gamma ray may lead to a large self-absorption correction. For example, the ^{232}Th foil of Table 1 has a maximum attenuation of 22 %, or an average correction of about 11 %, for the 537-keV transition. This represents an upper limit for the thickness of that foil.

Therefore, the self-attenuation of gamma rays, as well as the neutron self-shielding discussed later, will influence the foil selection.

NOTE 5—For other considerations in the selection of specific foils, see Guide E 844, Practice E 261, and Test Methods E 262, E 263, E 264, E 265, E 266, E 704, and E 705.

5. Apparatus

5.1 The gamma-ray detector should be a germanium-type detector (either Ge(Li) or intrinsic) with an energy resolution of 2.5 keV or better (full-width at half-maximum (FWHM) at 1173 keV). Associated equipment would include a multichannel pulse-height analyzer and a precision pulse generator with calibrated pulse-height and pulse-rate inputs into the detection system.

5.2 Foil and source holders should be used to provide precise positioning of a gamma-ray standard source and of each activated foil with respect to the detector. Required precision is about 0.2 mm or better in distance from the window of the detector or in lateral alignment.

5.3 National standard sources that are traceable to NIST (or their equivalent) should be used for calibration of the detection system.

6. Precautions

6.1 *Scattering Problems*—A sensor with a strong resonance absorption, such as a thick ^{235}U foil, should not be placed in front of a $1/v$ detector, and thick foils with covers should not be stacked because accurate corrections for the resultant scattering are difficult to determine. With an isotropic neutron-fluence, Φ_0 , incident on stacked foils, the reduction in the fluence rate caused by scattering at a given foil can be estimated by using the following equation:

$$\Phi = \Phi_0 e^{-\sum \sigma_i X_i}$$

where Φ is the attenuated fluence, \sum_i is a summation-over- i symbol, σ_i is the total macroscopic scattering cross section in cm^{-1} , and X_i is the thickness of the i th foil in centimetres. The summation is up to the foil of interest, located at its appropriate depth (distance from source) in the foil stack. For best results, the reduction in fluence rate should be less than 10 % for the foil located at the maximum depth.

6.2 *Foil Self-Shielding*—For the thicknesses of the foils recommended, the correction for self-shielding is appreciable only for pure gold foils (with its highly absorbing resonance at about 5 eV), the correction for a 0.025-mm thick foil being about a factor of two for epicalcium neutrons (that is, neutrons with energies greater than 0.5 eV) (15).

NOTE 6—Dilute aluminum-gold alloys are available and do not generally require self-shielding corrections.

6.3 *Fluence Nonuniformity*—If all the foils cannot be located in a region of uniform fluence rate (as determined by symmetry considerations), they can be located at different positions (and, hence, with different fluence rates) as long as the neutron energy spectrum is constant. If the fluence varies by more than 3 % from point to point, fluence monitors should be used with each foil. Around a Godiva-type reactor, sulfur foils can serve as monitors near the individual foils. Where

space is more limited, then nickel [$^{58}\text{Ni}(n,p)^{58}\text{Co}$], iron [$^{54}\text{Fe}(n,p)^{54}\text{Mn}$], or even aluminum [$^{27}\text{Al}(n,\alpha)^{24}\text{Na}$] should be considered for monitors. (See Practice E 261 for other relevant considerations.) Often a better solution is obtained by mounting all foils on a rotating disk or ring to ensure that they receive the same fluence.

6.4 Fluence Rate Depression—At low energies, fluence rate depression can be significant for bare thermal-neutron detectors near cadmium-covered foils if both are embedded in a moderator. This is because the cover on one foil can shadow adjacent bare foils. At high energies, depression can be significant for foils irradiated under the same conditions if the moderator contains reactor fuel. However, this is ordinarily not a problem, since in the sizable irradiation volumes normally used for radiation damage studies, the cadmium covers (as well as the foils) generally subtend a negligibly small solid angle at the point of any surrounding moderator or fuel. Fluence rate depression is usually insignificant for irradiation in a Godiva reactor glory hole.

7. General Handling Procedures

7.1 Foil Encapsulation—Fission foils should be encapsulated in sealed containers to avoid oxidation, loss of material, and for health-safety requirements. If a ^{239}Pu foil is used (instead of the much safer ^{235}U foil), it will require special encapsulation and periodic monitoring to check for leakage of the material. Copper encapsulation has been found satisfactory for ^{235}U , ^{238}U , ^{237}Np , and ^{232}Th foils. The thickness of the copper capsule should be about 0.1 to 0.25 mm at the flat surfaces and soldered at the periphery.

7.2 Foil Covers:

7.2.1 As noted in Table 1, cadmium covers of 0.5 to 1-mm thickness are prescribed for all fission foils and $1/v$ detectors. Cadmium covers also should be used for finite-threshold foils with trace impurities that yield the same reaction product by means of thermal-neutron capture. Examples are foils such as ^{238}U , ^{56}Fe , ^{58}Ni , and ^{27}Al with impurities of ^{235}U , ^{55}Mn , ^{59}Co , and ^{23}Na , respectively. Depending on the concentration, such impurities can lead to large correction factors. For large correction factors (that is, greater than 5 %), cadmium-covered foils made of the impurity materials should be irradiated. Then, corrections can be made with good accuracy if the impurity concentration in the primary threshold foil is accurately known. If the impurity concentration is not known, a thermal-neutron activation analysis of the foil can provide data for the necessary correction. Cadmium covers may not be required for foils irradiated in the empty “glory hole” of a fast-pulse reactor, a cavity in which little or no moderator material is normally present (that is, less than 0.5 g/cm^2).

7.2.2 Covers of ^{10}B for fission foils are useful when measuring a soft TRIGA spectrum. However, if a boral shield that provides good $4\text{-}\pi$ geometry surrounds the irradiation cavity, and if a negligible amount of moderator is contained within the shield, then the ^{10}B covers may not be required. The effect of the boral shielding should be accounted for properly when the neutron spectrum is adjusted with a proper computer code. More is said about boron cover corrections in 7.2.4.

NOTE 7—Spectra adjustment codes are discussed in Guides E 721 and E 944.

7.2.3 If no ^{10}B covers are used for the foils, and if the TRIGA irradiation cavity is only partially shielded by boral, then it will be difficult to determine the neutron spectrum from 10^{-2} MeV down to about $3 \times 10^{-7}\text{ MeV}$. If the TRIGA irradiation cavity has only partial boral shielding, it is important that all the fission foils, all the $1/v$ foils, and the foils with important $1/v$ impurities be placed in a boral “box” or a ^{10}B cover. For best results, a ^{10}B cover of 1 to 1.8 g/cm^2 of (93 %) ^{10}B should be used. In this way, the fraction of activations arising from neutrons in the energy range from $3 \times 10^{-7}\text{ MeV}$ to 10^{-2} MeV will be reduced greatly. The effect of the cover thickness can be accounted for by a spectrum adjustment code provided that the effective attenuation cross section that accounts for scattering in the cover is available. See 7.2.5.

7.2.4 For a Godiva-type reactor, ^{10}B covers may not be required, and cadmium covers may be sufficient for irradiation distances of less than 1 m from the reactor when the reactor is located a few metres above the concrete floor. Cadmium covers also may be used in the glory hole where the number of low-energy neutrons is insignificant. If ^{10}B covers are used, activities may require correction for scattering by the ^{10}B . The correction can be determined either experimentally with pure finite-threshold fission foils (^{237}Np or ^{232}Th) that contain negligible zero-threshold impurities, or with a neutron transport calculation that takes into account the thickness of the material.

7.2.5 The attenuation by a boron cover of the neutron fluence is not adequately treated by many of the spectrum adjustment codes (18). Some versions of the spectrum adjustment code, SAND II (11), for example, uses a simple exponential attenuation function versus energy, and because most irradiations are conducted in wide-beam or isotropic configurations, scattered neutrons are not in general lost from the beam. As a result, the absorption cross section of the boron should generally be used to determine the attenuation. However, in many configurations (such as narrow-beam geometry or down scattering of the neutrons to lower energy) the scattering portion of the cross section can remove additional neutrons and the true effective removal cross section value will fall somewhere in between the total and the absorption cross section. This is especially noticeable if the response of the foil is concentrated above the 10-keV limit where the B^{10} absorption ceases to dominate the cross section. Thus, for high-threshold fission foils such as ^{238}U and ^{237}Np or a normal threshold foil such as nickel, the additional scattering will result in additional attenuation. For example, some experiments and calculations indicate that these corrections are of the order of 10 % for a 1.65-g/cm^2 ^{10}B cover and a thin 12.7-mm diameter fission foil (12). Other work indicates that these scattering corrections may be somewhat larger (13). Strictly speaking, a calculation of the transport in the full-experiment geometry through the boron cover should be performed for each geometry (18). Measurements with a high-threshold foil, $\text{Ni}(n,p)^{58}\text{Co}$, have shown a transmission factor of 0.9 in a Godiva-type exposure geometry (23). This compares with a

calculated value (for which only the boron capture cross section is used) of 0.96.

NOTE 8—A monitor foil such as nickel used both inside and outside a boron ball can be used to normalize the boron-covered-fission-foil exposure to that of the rest of the foil set in case positioning errors are likely to be significant. The nickel ratio is not very sensitive to spectrum shape. The procedure is to multiply the fission foil activities by a factor that ensures that the ratio of nickel activities inside and outside the boron ball is about 0.9.

7.2.6 Another advantage of using covers (B, Cd) on broad energy-response foils is that it restricts that response and permits improved definition of the spectrum during the adjustment process. If both bare and Cd-covered resonance materials (such as Au and Na) are exposed, much better definition of the shape of the spectrum in the epithermal and thermal region can be obtained.

NOTE 9—Some versions of spectrum adjustment codes handle covers through the use of auxiliary codes that apply an energy-dependent-cover correction factor to the dosimetry cross section.

7.2.7 If the spectrum is to be well defined, then the foil set must contain a large fraction of the reactions from Table 1 and possess response functions spread as uniformly over energy as is possible. This is necessary to ensure that the spectrum adjustment codes can arrive at sufficiently restricted solutions. With broad response functions the calculated fluence at one energy can influence the calculated spectrum values at distant energies. If at all possible include ^{237}Np , and ^{239}Pu or ^{235}U to provide sensitivity between 10 keV and 1 MeV where few other reactions have significant response. Silicon devices are also sensitive in this energy region and can be used as spectrum sensors (14).

8. Certification of Foil Purity

8.1 The foil purity analysis results should be recorded permanently so that appropriate impurity corrections can be made. The acceptable uncertainty in the results mainly dictates what impurity concentrations are acceptable. It also depends on the nature and source of the neutron spectrum being measured (see 4.2). If, for example, the percentage impurity of ^{235}U in a foil of ^{238}U is known to be 400 ppm to an accuracy of 10 %, a separate ^{235}U foil can be irradiated in the same way as the primary foil to determine a proper correction factor. In this case, the impurity effect can be reduced to 10 % of its stated value (40 ppm) ^{235}U in ^{238}U by applying the correction factor. In determining the activity of a ^{238}U foil irradiated with a TRIGA spectrum to an uncertainty of 2 % or less, up to 2000 ppm of ^{235}U impurity could be tolerated (see 4.2).

9. Determination of Activities

9.1 A suitable set of sensors is placed in the neutron field under study. After irradiation, the specific activities of the sensor are determined by counting the gamma-ray emissions from each foil and applying appropriate corrections.

NOTE 10—Other energy response functions appropriate for spectrum

adjustment procedures measured by detection of other effects, such as emulsion tracks or even displacement damage, can also be used successfully. See Guide E 944, 4.1.

9.2 The measured specific activities of the activation foils are related to the incident neutron energy-fluence spectrum by the following equation:

$$R_j = \int_{\sigma}^{\infty} \sigma_j(E)\Phi(E) dE \quad 1 \leq j \leq n \quad (1)$$

where:

R_j = measured specific activity of an activated foil isotope j ,

$\sigma_j(E)$ = neutron cross section at energy E for isotope j ,

$\Phi(E)$ = incident neutron fluence differential in energy, and

n = number of reactions.

9.3 The differential neutron energy-fluence spectrum $\Phi(E)$ is calculated by means of a computer code that utilizes the specific activity data from the activation foil set. A number of these codes have been developed for this purpose and are available from the Oak Ridge National Laboratory Radiation Shielding Information Center (16).

10. Detector Calibration Procedures

10.1 Follow the general considerations in General Methods E 181 and Test Method E 265 on energy and efficiency calibration of the detector.

10.2 The germanium detector is usually operated at low temperatures (near the boiling point of liquid nitrogen). This requires the detector to be in a cryostat under vacuum. Normally, a thin window separates the detector's face from the outside environment. In such an enclosure, the exact position of the effective center of the active volume of the detector with respect to the cryostat window may not be known precisely.

10.3 Very low-activity foils must be placed close to the detector window in order to achieve a reasonable count rate. For such close foil-detector spacing, two problems occur that can affect the detector efficiency. One concerns the effects of finite source size on the effective detector solid angle, and the other concerns coincidence photon summing. Coincidence summing occurs when a radionuclide emits two or more cascade photons within the resolving time of the detector system. These problems are considered in the following sections that deal with determining detector efficiency.

10.3.1 Measure the count rate under each energy peak from a small diameter (about 2 mm) standard source at some specified distance, d (100 mm or greater), from the detector window. Use a long-lived mixed radionuclide standard source or several single radionuclide standard sources (or their equivalents) for these measurements (see Note 11). Determine the detector efficiency, $\epsilon(d)_s$, at this distance, d , from the source. The detector efficiency is defined as the ratio of the net count rate under the selected energy peak to the known gamma-ray emission rate of the standard source at that energy. A log-log plot of these data provides an efficiency-versus-energy curve

for later use in estimating the efficiency for foils of larger diameter than the point calibration source.

NOTE 11—An example of a mixed radionuclide standard source suitable for this purpose is NBS SRM 4275B.³ NBS is now NIST. An alternative method for calculating summing corrections is found in the documentation for this source. While this technique does not require the counting of foil materials in two locations, as discussed below, it does require that the detector's total efficiency curve be known. Experience has shown that a relatively crude knowledge of the total efficiency curve is sufficient to calculate summing corrections within a few percent for the foils in Table 1 except for ⁴⁸Ti.

10.3.2 To determine the detector efficiency for activated foil, *j*, select one of the higher-activity single-energy-transition foils, (for example, ¹⁹⁷Au(*n,γ*)¹⁹⁸Au with a 412-keV gamma ray), and measure the peak count rates at a position, *c*, close to the detector window and at the distant position, *d*. From the definition of detector efficiency, it can be seen for Foil *j* that the ratio of peak count rates is equal to the ratio of efficiencies at the respective positions as follows:

$$\frac{\dot{N}_p(c)_j}{\dot{N}_p(d)_j} = \frac{\epsilon(c)_j}{\epsilon(d)_j} \quad (2)$$

where \dot{N}_p in the net count rate under the selected energy peak. It is important to note that the count rate \dot{N}_p is actually defined as the average count rate during the count period,

$$\dot{N}_p = N_p/t_i,$$

where t_i is the count period.

10.3.3 Assume that the efficiency at position *d* is approximately the same for both the selected foil and the standard source. Then the efficiency for Foil *j* at the Position *c* is expressed as follows:

$$\epsilon(c)_j = \frac{\dot{N}_p(c)_j}{\dot{N}_p(d)_j} \epsilon(d)_s \quad (3)$$

where $\epsilon(d)_s$ is obtained from the calibration curve determined in 10.3.1.

10.3.4 Repeat the procedures of 10.3.2 and 10.3.3 for several other high-activity single-energy-transition foils with gamma-ray energies covering the energy range of interest. Use these data to determine an efficiency-versus-energy curve for a foil at the distance of *c* from the detector.

10.3.5 The procedures given in 10.3.1-10.3.3 are valid for either single-energy transitions or cascade transitions (two or more photons in coincidence). However, the efficiency-versus-energy curve determined in 10.3.4 for single-energy-transition foils is not applicable to cascade-transition foils because of the coincidence-summing effect. The efficiency for each cascade-transition radionuclide should be determined individually in order to avoid the uncertainties and efforts associated with calculating the summing corrections. The difference between the efficiency of a cascade-transition nuclide and the single-energy-transition efficiency curve at the same photon energy can be large for such cascade-transition nuclides as ⁶⁰Co, ⁵⁶Mn, ¹⁴⁰La, and ²⁴Na when these sources are counted very close to the face of the detector. For example, a difference

of 27 % has been observed for the 1596-keV gamma ray from ¹⁴⁰La when counting this source near the face of a 65-cm³ Ge(Li) detector (17).

10.3.6 Coincidence-summing effects are only important for very close detector-source geometries. Thus, the calibration efficiency curve determined with the standard source in 10.3.1 is valid even if there are cascade transitions from the standard source since the source-to-detector distance is 100 mm or greater. The additional uncertainty from summing for measurements with the source at 100 mm from the window of a 65-cm³ Ge(Li) detector is 1 % or less (17).

10.3.7 In order to get reasonable counting rates at both the close and distant locations relative to the detector, some planning of the counting procedures may be required when determining the detector efficiency for the various foils. At the distant location (100 mm or greater from the detector face), the count rate should be high enough to achieve good counting statistics in a reasonable time period. At the close location, the count rate should be low enough so that large and complex corrections for random summing and pulse pileup are avoided (see 12.1). One possible method for meeting these conflicting requirements is to use isotopes with reasonably short half-lives so that the foil can be counted first at the distant location and then a few days later at the close location when it has decayed to an appropriate level.

10.3.8 As an example of the positioning reproducibility required when foils are counted close to the detector window, suppose that the effective center of a particular detector is 21 mm from the front face of the cryostat window. Then, the location of the center of the source (either the standard or the activation foil) relative to the cryostat face must be reproduced within 0.2 mm for the uncertainty in the distance to the detector center to be within 2 %.

10.4 Another detector calibration procedure may be used for activation foils with half-lives longer than about 1 day. It utilizes fission neutron sources in irradiation facilities (for example, at the National Institute of Standards and Technology) where direct free-field neutron calibrations (18,19) are available. Such sources provide certified fluences of up to 10¹³ n/cm² in low-background environments.

10.4.1 The procedure involves irradiating a set of activation foils in calibrated neutron fields and then transporting the foils to the user's detector apparatus for counting. If the neutron fluence rate under investigation is similar to a fission spectrum, or if the detector response is energy independent over the energy range of interest, a direct neutron fluence transfer technique can be made. For details of this method, see 4.8.3 of Method E 261. The neutron fluence transfer technique relaxes the requirement to establish absolute detector efficiencies, and the uncertainties associated with absolute cross sections are significantly reduced because only ratios of the spectrum-averaged cross sections are required. It is important, however, for the spectrum-averaged cross section of the calibration source and of the reactor spectrum under consideration to be calculated from the same cross-section compilation.

10.5 If the activated foil has a decay scheme containing a significant number of low-energy gamma rays as well as the gamma ray of interest, insert a lead shield of an appropriate

³ NIST SRM 4275B available from Office of Standard Reference Materials, National Institute of Standards and Technology, Gaithersburg, MD 20899.

thickness between the foil and the detector. Choose a shield thickness that will significantly attenuate the low-energy gamma rays and avoid pulse pileup in the detector, but still allow a reasonable count rate for the desired gamma ray. For example, a lead shield of about 13-mm thickness is appropriate for counting some fission foils (^{237}Np , ^{235}U , ^{239}Pu , and ^{238}U) yielding the same fission product of interest, ^{140}La . If such a lead shield is required, then perform the calibration procedures of 10.3 or 10.4 with the lead shield in place and determine the detector efficiency for the resulting foil-shield-detector geometry. (For more details on ^{238}U and ^{237}Np , see Test Methods E 704 and E 705, respectively.)

11. Counting Procedures

11.1 Pulse-Height Analyzer Deadtime and Pulse Pileup:

11.1.1 Use of a precision pulse generator is recommended for determining the correction for the combined effects of multichannel analyzer deadtime and pulse-pileup losses. Use the pulser dynamically; that is, pulses from it are counted at the same time that the activated foils or standard source are measured. Adjust the pulser output (pulse amplitude) to place the peak in a low-background area of the spectrum being analyzed. Also, use a low repetition rate (about 60 Hz).

11.1.2 With a foil in the counting position and the pulser on, run the analyzer on “clock” time (as opposed to “live” time). The ratio of the number of pulses generated during the counting period to the number of counts recorded in the pulser peak in the analyzer gives the correction factor for the combined deadtime and pulse-pileup losses (20).

11.2 Background Corrections:

11.2.1 Minimize laboratory background counts by selecting a low-background counting area, by moving all radioactive sources other than the foil being counted from the counting area, and by placing a gamma shield around the detector. If the gamma shield is lead, it should be at least 50 mm thick.

11.2.2 Carefully determine the backgrounds for fission-foils since all such foils have some residual natural radioactivity and, also, because such foils often are reused due to their initial high cost. If the foils have been previously irradiated within a period of less than several half-lives of the gamma ray of interest, then measure the background at least twice. Allow sufficient time to elapse between the measurements so that gamma rays with relatively short half-lives can be distinguished from any long-lived components due to either natural radioactivity or to other fission fragments. Corrections can then be made for both short- and long-lived background components.

11.3 *Counting Redundancy*—It is recommended that each foil be counted on each of two or more calibrated counting systems. If there is disagreement by more than 5 %, repeat the count (and calibration if necessary). If only one counter is available, at least remove and replace the foil between readings.

12. Data Analysis

12.1 Peak-Area Analysis:

12.1.1 Use a consistent method of peak-area analysis for peaks originating from the precision pulser, the calibration source, and the activated foils.

12.1.2 In one basic method, plot the counts per channel around the peak. Subtract the baseline area (background) from the peak area by fitting a straight line through the baseline.

12.1.3 In counting fission foils, examine carefully the primary peak for the presence of a very close neighboring peak. If close neighboring peaks are present, use a peak-shape analysis technique. This analysis can be done either by hand or with a suitable computer code (for example, the SAMPO code (21)). Other peak analysis codes associated with commercial counting systems are also available. Good counting statistics are necessary for the peak-shape analysis to give reasonably accurate results. Accumulate at least 10 000 counts in the net peak area whenever possible.

12.2 *Peak Area Corrections*—Since in neutron effects testing of electronic parts, the usual interest is in the permanent damage from the integral fluence, the activities discussed in this section are determined in terms of detector counts (in the manner of Test Method E 265) rather than in terms of count rates. The derivations of the activities are given in the appendix. Correct the net peak areas determined in 12.1 for analyzer deadtime and pulse-pileup losses by multiplying by the correction factor discussed in 11.1.1. The efficiency of the counting system must be accounted for and correction made for self-absorption of the gamma rays by thick foils. Thus, the gamma-ray emission, N_γ , (number of photons emitted by the daughter isotopes caused by the fluence Φ in the foil) is given as follows:

$$N_{\gamma(tc)} = \frac{N_p(t_c)C_{pp}e^{[(\mu_a/\rho)(Z/2)]}}{\epsilon(c)} \quad (4)$$

where:

- N_p = net counts under the peak after counting for time t_c ,
- $\epsilon(c)$ = detector efficiency for the foil at Position c (counts per disintegration),
- C_{pp} = correction factor for analyzer deadtime and pulse pileup (ratio of pulses generated to counts in pulser peak),
- μ_a/ρ = mass absorption coefficient, cm^2/g , and
- z = combined thickness of the foil and any encapsulation material, g/cm^2 .

The exponential self-absorption correction is an approximation; however, it is reasonably accurate if the correction factor is less than 20 %.

12.3 Calculation of Sensor Activity:

12.3.1 To determine the activity of a sensor, correct for the decay of the activated sensor during the irradiation period, the waiting period, and the counting period. This requires the activity that would have existed if all of the fluence struck the foil in a time short compared to the half-life of the reaction of interest in the foil. This is because for fluence determination it is the total number of reactions that is needed. This activity is called A_o and is generally different from the activity A_i at the end of the irradiation. Thus, for a steady-state irradiation at a constant fluence rate, the foil activity for fluence determination is given as follows:

$$A_o = \frac{N_\gamma(t_c)\lambda^2 t_i e^{\lambda t_w}}{P_\gamma Y_f (1 - e^{-\lambda t_c})(1 - e^{-\lambda t_i})} \quad (5)$$

where:

- A_o = activity of the foil, that is, disintegration rate of neutron-activated nuclide of interest, corrected for decay during the irradiation, number per second,
- N_γ = gamma-ray emission (photons emitted during t_c from Eq 4),
- λ = decay constant of product nuclide ($= 0.693/\text{half-life}$),
- P_γ = probability per nuclear disintegration for the emission of the gamma ray being counted,
- Y_f = fractional fission yield for the product of interest (for fission foils),
- t_w = waiting time between end of irradiation period and beginning of counting period,
- t_c = counting period, and
- t_i = irradiation period.

12.3.2 Additional corrections to Eq 3 are required if significant neutron self-shielding or fluence depression occurs during irradiation. The value of A_o can be found from the activity at the end of the irradiation, A_i , by multiplying A_i by $\lambda t_i / (1 - e^{-\lambda t_i})$ provided the fluence rate during t_i was a constant. For short irradiation times, $t_i \ll 1/\lambda$, the correction approaches 1.0 and A_i approaches A_o . See the appendix for the derivations.

12.3.3 The absolute-specific-decay-corrected activity of a foil of Isotope j is given by the following equation:

$$R_j = \frac{A_o}{N_o} \quad (6)$$

where:

- R_j = specific activity of Isotope j (disintegrations per second per atom available for activation) assuming correction has been made for decay during t_i , and
- N_o = number of atoms of Isotope j in the foil available for activation.

The factor N_o can be expressed as follows:

$$N_o = \frac{N_A f m}{M} \quad (7)$$

where:

- N_A = Avogadro's number, 6.0225×10^{23} ,
- f = fractional mass abundance of Isotope j in the foil,
- m = total mass of the foil, and
- M = atomic mass of activated nuclide.

12.3.4 The specific activities of all the activation foils are used as input data for an appropriate computer code to determine the energy-fluence spectrum of the neutron source. See Guide E 721.

13. Precision and Bias

13.1 The factors that determine the uncertainty of the measured sensor specific activities are as follows:

13.1.1 Counting statistics,

13.1.2 Reproducibility of the location of the standard source or foil with respect to the detector,

13.1.3 Reproducibility in the determination of net full-energy peak counts (peak-area analysis, background subtraction, and coincidence photon summing correction),

13.1.4 Systematic uncertainties, which are considered separately in 13.3,

13.1.5 Error associated with the positioning of foils in a nonuniform field,

13.1.6 Monitor normalization between separate runs, and

13.1.7 Uncertainties in the cross sections of the reactions and of the covers.

13.2 For example, a typical foil-counting system could have magnitudes of random errors (each equaling one standard deviation, σ) as follows:

Source of Random Error	Uncertainties % (Non-Fission Foil)	Uncertainties % (Fission Foil)
Counting statistics, σ	1.0	1.0
Source or foil location, σ	1.0	2.0
Peak counts, σ	1.0	2.0
Total, σ_r	1.7	3.0

The total random error, σ_r , in this example is obtained by combining the individual values in quadrature (that is, the square root of the sum of the squares).

13.3 Sources of systematic (nonrandom) uncertainties (along with examples of typical estimated values of these uncertainties) are listed as follows:

Source of Systematic Error	Uncertainties % (Non-Fission Foil)	Uncertainties % (Fission Foil)
Calibration of gamma detector	3.0	3.0
Correction for finite source size	0.5	0.5
Correction for analyzer deadtime and pileup	1.0	1.0
Correction for self-shielding and fluence depression	2.0	2.0
Nuclear data:		
Branching ratios of gamma ray scheme	2.0	2.0
Fission yields	...	3.0
Half-lives	1.0	1.0
Total, errors combined in quadrature	4.3	5.2

13.4 An estimate of the overall uncertainty in determining the specific activity of a foil is obtained by combining the random and systematic uncertainties in quadrature. Thus, for the example in 13.2 and 13.3, the overall uncertainty for nonfission foils is $[(1.7)^2 + (4.3)^2]^{0.5} = 4.6\%$, and for the fission foils it is $[(3.0)^2 + (5.2)^2]^{0.5} = 6.0\%$. Of course, these values are meant to approximate the errors in counting a typical foil and would not necessarily be representative of any specific case.

13.5 Perform a detailed error analysis for each foil measured and give details in a report of the results.

14. Keywords

14.1 activation foils; neutron activation; neutron spectra; radiation-hardness testing

APPENDIX
(Nonmandatory Information)
X1. DETERMINATION OF ACTIVITIES FROM COUNTING DATA
X1.1 Activity at the End of the Irradiation Period:

X1.1.1 The first objective is to determine the activity at the end of the irradiation, A_i , from counts recorded during the count period, t_c , which starts a time t_w after the end of the irradiation period. Note that this relationship does not depend on the irradiation history before t_i . All that is sought here is the emission rate caused by the daughter isotopes, N_d , that are present at t_i . Their decay rate is λN_d , and it does not matter how the quantity N_d came to be.

X1.1.2 The activity at the beginning of the count period is as follows:

$$A(t_w) = A_i e^{-\lambda t_w} \quad (\text{X1.1})$$

Then the number of gammas emitted during the count time, t_c , is as follows:

$$\begin{aligned} N_\gamma(t_w, t_c) &= A_i e^{-\lambda t_w} \int_0^{t_c} e^{-\lambda t} dt, \\ &= \frac{A_i}{\lambda} e^{-\lambda t_w} (1 - e^{-\lambda t_c}), \end{aligned}$$

which leads to

$$A_i = \frac{N_\gamma \lambda e^{\lambda t_w}}{1 - e^{-\lambda t_c}} \quad (\text{X1.2})$$

Notice that this relationship does not depend on the time history of the fluence or on t_i . Another relationship must be found to connect this activity to the activity A_o (Eq 3), and the fluence, Φ . But before that is done, the saturated activity, A_s , can be established. This is the activity that would exist after a steady-state irradiation for a time $t_i \gg 1/\lambda$.

X1.2 Saturated Activity:

X1.2.1 The total rate of change of daughter isotopes during irradiation is as follows:

$$\frac{dN_d(\text{total})}{dt} = \sigma \Phi N - \lambda N_d(\text{total}) \quad (\text{X1.3})$$

where:

σ = spectrum averaged cross section for the foil reaction,

Φ = neutron fluence rate, and

N = number of target nuclei in the foil.

These radioactive nuclei are being created at a rate $\sigma \Phi N$ and decay at a rate $\lambda N_d(\text{total})$. Only the second term on the right is responsible for the gamma rays that are to be counted.

X1.2.2 The solution to Eq X1.3 for the number of daughter nuclei present at the end of a steady-state irradiation at rate ϕ starting at $t = 0$ and ending at $t = t_i$ is as follows:

$$N_d(\text{total}, t_i) = \frac{\sigma \Phi N}{\lambda} (1 - e^{-\lambda t_i}) \quad (\text{X1.4})$$

The decay rate is then $\lambda N_d(\text{total}, t_i)$, the second term on the right of Eq X1.3.

X1.2.3 The decay rate is related to the gamma-ray emission rate needed to calculate the activity by the relation as follows:

$$A_i = \lambda P_\gamma Y_f N_d(\text{total}, t_i) \quad (\text{X1.5})$$

where P_γ and Y_f are the probabilities defined in Eq 5. (From this direction of derivation (from fluence rate back to activity at t_i), the activity does depend on t_i .) After a period of irradiation long compared to $1/\lambda$, the decay rate will equal the creation rate, from Eq X1.3 (where $dN_d(\text{total})/dt = 0$), as follows:

$$P_\gamma Y_f (\lambda N_d = \sigma \Phi N) = A_s$$

Thus

$$A_s = \frac{A_i}{(1 - e^{-\lambda t_i})}$$

Again the condition was that A_i was established by a steady-state irradiation for a period t_i . The conversion factor from A_i to A_s is very large if the irradiation time is $\ll 1/\lambda$.

X1.3 Relationship of A_o and A_i :

X1.3.1 Since the activity is proportional to the number of daughter isotopes, the relationship between A_o (the activity that would have been induced by a Φ deposited during a $t_i \ll 1/\lambda$) and A_i (the activity at the end of the irradiation at t_i by comparing the number of daughter isotopes created by Φ with those remaining at the end of the irradiation at t_i). This is because the multiplying factors ($\lambda P_\gamma Y_f$) are the same.

$$N_{do} = \sigma \Phi N = \text{total daughters generated} \quad (\text{X1.6})$$

From Eq X1.4 calculate as follows for a steady fluence rate:

$$\begin{aligned} N_d(\text{total}, t_i) &= \frac{\sigma \Phi N}{\lambda} (1 - e^{-\lambda t_i}) \\ &= \frac{\sigma \Phi N}{\lambda t_i} (1 - e^{-\lambda t_i}) \end{aligned}$$

Therefore

$$\frac{N_{do}}{N_d(\text{total}, t_i)} = \frac{A_o}{A_i} = \frac{\lambda t_i}{(1 - e^{-\lambda t_i})} \quad (\text{X1.7})$$

For short pulses $A_o = A_i$. It is important to note here that N_{do} is not really related to the N_d in Eq X1.5 even though they look similar. N_{do} is the number of daughter isotopes created by the fluence Φ . N_d in Eq X1.5 is the equilibrium population at constant fluence rate Φ .

X1.3.2 The procedure for finding the activity, A_o , to determine the fluence in the manner of Test Method E 265, is to find the activity at the end of the irradiation with Eq X1.2 and then multiply it by the correction factor in Eq X1.7. This yields Eq

5. This is also the same activity used in determining a spectrum from a set of foils as is discussed in Guide E 721.

REFERENCES

- (1) Rose, P. F., ed., “ENDF-201, ENDF/B-VI Summary Documentation,” *Brookhaven National Laboratory Report BNL-NCS-1741*, 4th Edition, October 1991. The cross-section libraries are distributed by the National Nuclear Data Center, Brookhaven National Laboratory. The file format is described in Ref 3.
- (2) Rose, P. F., and Dunford, C. L., ed., “ENDF-102, Data Formats and Procedures for the Evaluated Nuclear Data File ENDF-6,” *Brookhaven National Laboratory Report BNL-NCS-44945-01/04 Rev*, July 1990, revised April 2001.
- (3) “International Reactor Dosimetry File (IRDF-90),” assembled by N. P. Kocherov, et al, International Atomic Energy Agency, Nuclear Data Section, IAEA-NDS-141 Rev. 0, October 1993. This dosimetry library is available within the United States by contacting the National Nuclear Data Center at Brookhaven National Laboratory.
- (4) Sallee, W., Flanders, M., Sparks, M., and Meason, J., “Hog Ring Revisited—A New Look at the WSMR-FBR Neutron Spectrum,” Appendix B, TR-NEL-017-86, White Sands Missile Range, NM, November 1986.
- (5) Kelly, J. G., and Vehar, D. W., “Measurement of Neutron Spectra in Varied Environments by the Foil-Activation Method With Arbitrary Trials,” SAND87-1330-VC-2, Sandia National Laboratories, Albuquerque, NM, December 1987.
- (6) Grundl, J., and Eisenhauer, C. M., eds., “Compendium of Benchmark Neutron Fields for Reactor Dosimetry,” 85-3151, *National Bureau of Standards Internal Report*, 1985.
- (7) England, T., and Rider, B. F., “Evaluation and Compilation of Fission Product Yields 1991,” *Brookhaven National Laboratory Report Number ENDF-349, LA-UR-94-3106, October 1994*. The ENDF/B-VI Fission Product Yield Data Tapes are distributed by the National Nuclear Data Center, Brookhaven National Laboratory. The file format is described in Ref 3. The report is available from <http://t2.lanl.gov/publications/yields/endl349>.
- (8) Mannhart, W., “The Discrepancy Between Differential & Integral Data on Ti-47(n,p),” *Nuclear Data for Basic and Applied Science*, Vol 1, *Proceedings of the International Conference*, Santa Fe, NM, May 13–17, 1985. Editors P. G. Young, R. E. Brown, G. F. Auckampanah, P. W. Tisowski, L. Stewart: Gordon and Breach, publisher, 1986.
- (9) Fu, C. Y., and Hetrick, D. M., “Experience in Using the Covariances of Some ENDF/B-V Dosimetry Cross Sections: Proposed Improvements and Addition of Cross-Reaction Covariance,” *Proceedings of the Fourth ASTM-EURATOM Symposium on Reactor Dosimetry: Radiation Metrology Techniques, Data Bases, and Standardization*, Vol II, National Bureau of Standards, Gaithersburg, Maryland, March 22–26, 1982, pp. 877–887, Report Number NUREG/CP-0029, CONF-820321/V2. The GLUCS library currently contains 14 dosimetry reactions and is available from D. Hetrick at Oak Ridge National Laboratory.
- (10) Jagdish, K. Tuli, *Nuclear Wallet Cards*, July 1990, National Nuclear Data Center, Brookhaven National Laboratory, Upton, New York.
- (11) Berg, S., and McElroy, W. N., *A Computer-Automated Iterative Method for Neutron Flux Spectra Determination by Foil Activation*, AFWL-TR-67-41, Atomics International, September 1967 (used for thermal spectrum).
- (12) Harvey, J. T., Meason, J. L., and Wright, H. L., “Neutron Attenuation Due to Boron-10 Shields,” *Nuclear Science and Engineering*, Vol 67, 1978, pp. 343–344.
- (13) Lowrey, A. R., Friddel, K. D., and Milliman, L. D., “Fast Burst Reactor Neutron Damage in Silicon Transistors,” *IEEE Transactions on Nuclear Science*, NS-23, No. 6, 1976, pp. 2013–2019.
- (14) Kelly, J. G., and Griffin, P. J., “Comparison of Measured Silicon Damage Ratios with ASTM E722-85 and NJOY Calculated Damage,” *Proceedings of the Seventh ASTM-Euratom Symposium on Reactor Dosimetry*, Strasbourg, France, Aug. 27–31, 1990, pp. 711–718.
- (15) Sanford, G. S., and Secksinger, J. H., “Thickness Corrections for Neutron-Activated Gold Foils,” *ANL-7545*, Argonne National Laboratory, Argonne, IL 60439, 1969.
- (16) Roussin, R. W., ed., “RSIC Computer Code & Data Collections: A Capsule Review of the Computer Code Collection (CCC) Peripheral Shielding Routines (PSR) and Data Library Collection (DLC) Packaged by the Radiation Shielding Information Center,” Oak Ridge National Laboratory Radiation Shielding Information Center catalog, October 1990.
- (17) Debertin, K., and Schötzig, U., “Coincidence Summing Corrections in Ge(Li)-Spectrometry at Low Source-to-Detector Distances,” *Nuclear Instruments and Methods*, No. 158, 1979, pp. 471–477.
- (18) Kazi, A. H., McGarry, E. D., and Gilliam, D. M., “Standardization of Fast Pulse Reactor Dosimetry,” *Neutron Standards and Application*, NBS SP 493, 1977, pp. 335–341.
- (19) McGarry, E. D., Kazi, A. H., Davis, G. S., and Gilliam, D. M., “Absolute Neutron-Flux Measurement at Fast Pulse Reactors with Calibration Against Californium-252,” *IEEE Transactions on Nuclear Science*, NS-23, No. 6, 1976, pp. 2002–2006.
- (20) Cohen, E. J., “Live Time and Pile-Up Corrections for Multichannel Analyzer Spectra,” *Nuclear Instruments and Methods*, No. 121, 1974, pp. 25–32.
- (21) Routti, J. T., “SAMPO, A Fortran IV Program for Computer Analysis of Gamma Spectra from Ge(Li) Detectors, and Other Spectra with Peaks,” *UCRL-19452*, 1969.
- (22) Griffin, P.J., Kelly, J.G., “A Rigorous Treatment of Shielding and Covers in Neutron Spectra Determinations,” *IEEE Transactions on Nuclear Science*, NS-42, No. 6, 1995, pp. 1878–1885.
- (23) Evaluated Nuclear Structure Data File (ENSDF), a computer file of evaluated nuclear structure and radioactive decay data, that is maintained by the National Nuclear Data Center (NNDC), Brookhaven National Laboratory (BNL), on behalf of the International Network for Nuclear Structure Data Evaluation, which functions under the auspices of the Nuclear Data Section of the International Atomic Energy Agency (IAEA). The URL is <http://www.nndc.bnl.gov/nndc/ensdf>. The detailed retrieval option must be selected to get the data quoted here. The data quoted here comes from the database as of January 1, 2002.
- (24) Griffin, P.J., Williams, J.G., “Least Squares Analysis of Fission Neutron Standard Fields” *IEEE Transactions on Nuclear Science*, NS-44, No. 6, 1997, pp. 2071–2078.
- (25) E. Browne, R. Firestone, *Table of Radioactive Isotopes*, John Wiley & Sons, 1986.
- (26) H.D. Lemmel, “X-Ray and Gamma-Ray Standards for Detector Calibration,” International Atomic Energy Agency, IAEA-TECDOC-610, September 1991.
- (27) EPJ. Griffin, J.G. Kelly, T.F. Luera, J. VanDenburg, SNLRML Recommended Dosimetry Cross Section Compendium, Sandia National Laboratories, Albuquerque, NM, report SAND92-0094, November 1993.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).