



Designation: F 1638 – 95 (Reapproved 1999)

Standard Terminology for Surgical Tissue/Dressing/Pick-Up Forceps (Thumb-Type)¹

This standard is issued under the fixed designation F 1638; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This terminology defines basic terms and considerations for components of thumb-type surgical forceps. Instruments with this terminology are limited to those fabricated from stainless steel and used for surgical procedures.

2. Referenced Documents

2.1 ASTM Standards:

F 899 Specification for Stainless Steel Billet, Bar, and Wire for Surgical Instruments²

F 921 Definitions of Terms Relating to Hemostatic Forceps²

3. Terminology

3.1 Definitions

atraumatic—teeth that would interdigitate except for being

spaced apart a predesigned distance so they will not stress, crush, or otherwise traumatize the tissue being grasped.

guide pin—a pin affixed to the inside of one of the forceps halves that aligns with a hole on the other tweezer half without protruding through when closed.

guide pin hole—the hole in one forceps half into which the guide pin fits without passing through when closed.

mesh—an alignment of opposing teeth. The teeth may be in-line or angled.

mouse teeth—distal tip teeth that interdigitate.

scissoring—lateral misalignment.

set—the at rest position of the instrument halves that will provide the intended closing relationship of fit and force.

stop pin—a pin of preset length affixed to the inside of one of the tweezer halves designed to limit teeth contact upon closure and prevent their damage.

teeth—serrations formed on the inside faces of the distal end of the tweezer halves.

tissue forceps—a device formed in two generally symmetrical halves with their proximal ends secured together and set so their distal ends will stay separated unless pressed together.

¹ This terminology is under the jurisdiction of ASTM Committee F04 on Medical and Surgical Materials and Devices and is the direct responsibility of Subcommittee F04.33 on Medical/Surgical Instruments.

Current edition approved Oct. 10, 1995. Published December 1995.

² *Annual Book of ASTM Standards*, Vol 13.01.



APPENDIX

(Nonmandatory Information)

X1. RATIONALE

X1.1 Because there is a clinical need for a variety of instruments for surgical procedures, they are manufactured in various configurations and from various types of stainless steel. For practical purposes and patient safety, these devices supplied by different manufacturers necessitate a defined system of terms.

X1.2 The terms defined in this terminology are the most commonly used for this type of forceps. However, the intent is not to prohibit technological innovation or to exclude instruments manufactured with other types of features.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org).